


LABORATORY CATALOGUE

for milk analysis







Dear ladies and gentlemen,

The name Funke-Gerber stands for innovative chemical milk analysis combined with quality, continuity and reliability. Thousands of instruments installed all over the world, used daily by laboratory specialists, confirm our excellent reputation, a reputation that has been strengthened by longtime trust and cooperation between us and our business partners. Furthermore, Funke-Gerber is considered to be a reliable business partner and supplier of laboratory equipment who offers a very good price-performance ratio. With pride and satisfaction we look back on our more than century-long performance.

In this newly revised catalogue, I would in particular like to point out our new products "LactoFlash" and "LactoStar". Of course, all other equipment is also up-to-date. As you know from previous catalogues, editorial contributions make up a large part of our publications. We are happy that we have been able to expand this important section with additional contributions from highly competent authors.

Our standard product range covers the entire range of chemical milk analysis. Should you have any special requirements that go beyond our product range, please do not hesitate to approach us with your respective inquiries. We will promptly respond by submitting an attractive quote.

We look forward to working with you!

K. Schaefer, CEO

CONTENTS

Preface	3
Funke - Dr. N. Gerber, tradition, progress, continuity	6
Sampling and sampling preparation	9
Butyrometric determination of fat content according to Dr. N. Gerber's method	13
<i>A detailed description by graduate chemist Alfred Toepel</i>	
Butyrometric determination of the fat content of various dairy products	20
<i>Determination of fat content in cream, ice cream, cheese, butter, milk powder, etc.</i>	
Determination of fat content (butyrometer)	26
<i>The butyrometer - an overview of the entire product spectrum</i>	
Determination of fat content (accessories)	30
<i>Implements and utensils used in the determination of fat content</i>	
LactoStar - LactoFlash	36
Centrifuge accessories	46
The Nova Safety bench centrifuge	47
Milk centrifuges	48
<i>Some important points about the acquisition and operation of a Gerber centrifuge K. Schaefer, graduate engineer, reports</i>	
The SuperVario-N multipurpose centrifuge	52
Water baths	53
Kjeldahl's nitrogen determination method	56
<i>Anna Politis, graduate engineer, reports</i>	
Equipment and accessories for Kjeldahl's protein determination method	61
pH measurement, operation and maintenance of the pH meter	63
Equipment and accessories for the measurement of pH value	65
Titration instruments / determination of acidity	68
Sediment tester	70
<i>Filter papers, Sedilab, Aspilac, etc.</i>	
Estimation of bacteria count	71
General laboratory equipment	72
<i>Butter melting beaker, testing spoon, spatula, aluminium foil, crystalline quartz sand, Bunsen burner, separating funnel, thin layer chromatography, etc.</i>	
Refractometers	75
Moisture measurement	76
Laboratory scales	79
Heating cabinets, incubators, refrigerated incubators, laboratory furnaces	80
Viscosity measurement, inhibitor detection	81
Densitometry / hydrometers	82
Thermometers, moisture measuring devices	86
Freezing point measurement	90
<i>One of the main focuses of Funke - Dr. N. Gerber Labortechnik GmbH K. Schaefer, graduate engineer, W. Spindler, graduate physician</i>	
CryoStar I, CryoStar^{automatic} and accessories	96

Solubility index mixer, jolting volumeter	100
Heating and mastitis detection	101
Laboratory utensils	102
Laboratory equipment	106
<i>ColonyStar bacteria counter, autoclaves, incubators, magnetic stirrers, photometers, microscopes, water distillation apparatuses, water baths</i>	
The use of reference material in the laboratory	112
<i>Dr. Ulrich Leist, DRRR GmbH, reports</i>	
Laboratory glassware	120
Laboratory appliances	126
Alphabetical index	132

TRADITION PROGRESS CONTINUITY



Funke - Dr. N. Gerber Labortechnik GmbH

Partners in dairy farming since 1904

Since 1904, Funke-Gerber has been an important player in dairy farming, both at home and abroad. The production of laboratory equipment for the testing of milk and foodstuffs is among its crowning achievements.

The manufacture of centrifuges together with butyrometers and other appliances for the determination of fat content according to Dr. N. Gerber's method continues to be central to our business activity. Above and beyond this classical field, the company develops and produces modern electronic devices for milk analysis.

The appliances of the **"CyroStar"** line for the determination of freezing points are highly regarded on account of their precision and reliability and have been in use in many dairies and institutes for years.

A new era in routine laboratory analysis has begun with the new **"LactoStar"** and **"LactoFlash"** appliances.

Know how and continuous development make Funke-Gerber an important player in dairy farming.

Decades of cooperation and trust have given our company, in association with our numerous business partners who represent Funke-Gerber in their countries, the necessary global presence to ensure the provision of products to our customers.

Since 1904, the name Funke-Gerber has stood for quality, reliability and continuity.

PRODUCTS:

The company develops, manufactures and markets the following equipment and accessories worldwide:

- All equipment and accessories for the „Gerber method of determining fat content“: **centrifuges, water baths, reading lamps, butyrometers**
- „CryoStar“ freezing point determination units
- „LactoStar“ and „LactoFlash“ milk analysis devices
- pH meters
- General laboratory equipment

ACTIVITIES:

Turnkey installation or the design of complete laboratories in the following special fields:

- The milk-processing industry
- Dairies, milk collection centres
- Cheese dairies, butter works, ice-cream, condensed milk and powdered milk factories

Company profile:

Founded: 1904

Managing director:

Graduate engineer Konrad Schaefer

Authorised signatory:

graduate economist Georg Hoernle

Address:

Funke-Dr.N.Gerber Labortechnik GmbH

Ringstraße 42

12105 Berlin

Telephone: (+49-30) 702 006-0



Fax: (+49-30) 702 006-66

E-mail: kontakt@funke-gerber.de

Website: www.funke-gerber.de

Milk sampler

stainless steel,
with valve for automatic drainage



3000	1 ml	$l_3 = 375 \text{ mm}, l_2 = 343 \text{ mm}, l_1 = 32 \text{ mm}, b_1 = 31 \text{ mm}, d_1 = 10 \text{ mm}$
3001	2 ml	$l_3 = 405 \text{ mm}, l_2 = 355 \text{ mm}, l_1 = 50 \text{ mm}, b_1 = 35 \text{ mm}, d_1 = 10 \text{ mm}$
3003	5 ml	$l_3 = 290 \text{ mm}, l_2 = 235 \text{ mm}, l_1 = 55 \text{ mm}, b_1 = 31 \text{ mm}, d_1 = 14 \text{ mm}$
3004	10 ml	$l_3 = 305 \text{ mm}, l_2 = 235 \text{ mm}, l_1 = 70 \text{ mm}, b_1 = 31 \text{ mm}, d_1 = 18 \text{ mm}$
3007	20 ml	$l_3 = 315 \text{ mm}, l_2 = 240 \text{ mm}, l_1 = 75 \text{ mm}, b_1 = 35 \text{ mm}, d_1 = 30 \text{ mm}$
3008	40 ml	$l_3 = 335 \text{ mm}, l_2 = 235 \text{ mm}, l_1 = 100 \text{ mm}, b_1 = 32 \text{ mm}, d_1 = 28 \text{ mm}$
3010	50 ml	$l_3 = 365 \text{ mm}, l_2 = 240 \text{ mm}, l_1 = 125 \text{ mm}, b_1 = 32 \text{ mm}, d_1 = 28 \text{ mm}$
3011	100 ml	$l_3 = 370 \text{ mm}, l_2 = 235 \text{ mm}, l_1 = 130 \text{ mm}, b_1 = 32 \text{ mm}, d_1 = 38 \text{ mm}$

Milk stirrer

stainless steel, perforated disk,
Ø 160 mm, 770 mm long


3021



Dipper

aluminium with spout, handle approx. 50 cm long


3030	125 ml	$l_3 = 625 \text{ mm}, l_2 = 540 \text{ mm}, l_1 = 85 \text{ mm}, b_1 = 53 \text{ mm}, d_1 = 43 \text{ mm}$
3031	250 ml	$l_3 = 620 \text{ mm}, l_2 = 540 \text{ mm}, l_1 = 80 \text{ mm}, b_1 = 53 \text{ mm}, d_1 = 65 \text{ mm}$



Scoop

stainless steel

3033	130 ml	$l = 350 \text{ mm}, \text{inner scoop } \varnothing = 79 \text{ mm}$
3034	250 ml	$l = 465 \text{ mm}, \text{inner scoop } \varnothing = 97 \text{ mm}$
3035	450 ml	$l = 480 \text{ mm}, \text{inner scoop } \varnothing = 118 \text{ mm}$



Milk sample bottle

80 ml, PE without metal bottom
(e.g. for art. no. 3510, 3530)
(for cap see art. no. 3043)

3040



Milk sample bottle

3041 50 mL, PP with metal bottom (e.g. for art. no. 3510, 3530)

3042 Stopper with groove (for art. no. 3041)

3043 Cap (for art. no. 3040)

Rubber stopper

for special solubility index tubes art. no. 3637

3050 19 x 24 x 25 mm

Cleaning brush

(for art. no. 3040, 3041, 3637)

3080 length: 300 mm



Wire cradle

plastic-coated wire, for 50 bottles,

3091 each 50 ml (for art. no. 3041)





Cheese trier

chrome-nickel steel, with plastic handle

3120 $l_1 = 125 \text{ mm}$, $l_2 = 60 \text{ mm}$, $l_3 = 190 \text{ mm}$, $b_1 = 85 \text{ mm}$, $b_2 = 13 \text{ mm}$

3121 $l_1 = 140 \text{ mm}$, $l_2 = 48 \text{ mm}$, $l_3 = 205 \text{ mm}$, $b_1 = 80 \text{ mm}$, $b_2 = 19 \text{ mm}$

3122 $l_1 = 150 \text{ mm}$, $l_2 = 75 \text{ mm}$, $l_3 = 225 \text{ mm}$, $b_1 = 80 \text{ mm}$, $b_2 = 21.5 \text{ mm}$



Cheese trier

100% stainless steel

3124

$l_1 = 125 \text{ mm}$, $l_2 = 40 \text{ mm}$, $l_3 = 165 \text{ mm}$, $b_1 = 65 \text{ mm}$, $b_2 = 15 \text{ mm}$



Milk powder collector

stainless steel for approx. 230 ml,

3125

exterior approx. $\varnothing = 28 \text{ mm}$, fill length = 375 mm



Butter trier


chrome-nickel steel, with metal handle

3130

$l_3 = 343 \text{ mm}$, $l_2 = 73 \text{ mm}$, $l_1 = 255 \text{ mm}$, $b_1 = 82.5 \text{ mm}$, $b_2 = 23 \text{ mm}$

3131

$l_3 = 410 \text{ mm}$, $l_2 = 75 \text{ mm}$, $l_1 = 320 \text{ mm}$, $b_1 = 80 \text{ mm}$, $b_2 = 22 \text{ mm}$



BagMixer 400 with window

capacity: 80 – 400 ml, 230 V/50 Hz

3139

17 kg, 400 x 270 x 260 mm

BagMixer 400 without window

capacity: 80 – 400 ml, 230 V/50 Hz

3140

17 kg, 400 x 270 x 260 mm



Accessories for BagMixer 400

3141

Disposable plastic bag 400 ml, sterile

3142

Filter bag, 400 ml, sterile

3143


Bag clasps

3144

Stand for 10 bags

BUTYROMETRIC DETERMINATION OF FAT CONTENT ACCORDING TO GERBER'S METHOD

by graduate chemist Alfred Toepel





Graduate chemist Alfred Toepel started lecturing at the School of Dairy Farming in Halberstadt in 1960. In 1992 he took over the Department of Training in Oranienburg.

He is also author of the textbook The Chemistry and Physics of Milk.

The butyrometric determination of fat content in milk was developed by **Dr. N. Gerber** in 1892 and incorporated into official regulations as a sulphuric acid process in 1935. The rapid testing method appears both in German standards (e.g. DIN 10479) and international standards (e.g. ISO 2446).

Determination of fat content according to Gerber is a rapid testing method and is still used today despite the introduction of automated methods for determining fat content in dairy laboratories. The advantages of the Gerber method over modern rapid testing methods are:

- Omission of the need for time-consuming calibration of the measuring gauge;
- Relatively low investment costs and hence low costs in performing quick tests on individual samples;
- It can be used on all types of milk.

The disadvantage is the use of very corrosive, concentrated sulphuric acid, which necessitates the observation of special precautions and the disposal of the sulphuric acid mixture in an environmentally suitable way.

PRINCIPLES OF THE METHOD

The determination of fat content according to Gerber involves running off the fat into a special measuring vessel, the butyrometer, and determining its volume as a percentage by mass. The fat is present in the milk in the form of small globules of various diameters, from 0.1 to 10 micrometers. The globules of fat form a consistent emulsion with the milk liquid. All globules of fat are surrounded by a protective coating, a fat globule membrane which is made up of phospholipids, a fat globule coat protein and hydrate water. This protein coating the fat globules prevents them from coalescing and stabilizes the emulsified state.

In order to completely isolate the fat, the protection coating around the fat globules must be destroyed. This is done with concentrated sulphuric acid of 90-91 % by mass. The sulphuric acid oxidizes and hydrolyzes the organic components in the protective coating around the fat globules, the lactoprotein fractions and the lactose. This produces a high heat of reaction in addition to the heat of dissolution. The butyrometer gets quite hot. The oxidation products turn the resulting solution brown. The released fat is then isolated by centrifuging, whereby the addition of amyl alcohol facilitates phase separation and a sharp delineation is produced between the fat and the acid solution. The fat content of the milk can be read off as a mass percent content on the butyrometer scale.

APPLICATION

This process can be used for untreated and pasteurized milk with a fat content of 0-16 % for milk which contains a suitable preservative as well as for homogenized milk.

CHEMICALS NEEDED

Sulphuric acid, H_2SO_4

Requirements:	Hazard symbol:	Hazard rating:
Density at 20°C (1,818 ± 0,003) g per ml ⁻¹		C2 R 35
• colourless or only slightly discoloured and free from any substances which might influence the outcome		S 2 - 26 - 30

Please note:

The required density corresponds to 90 to 91 % by mass. Stronger or weaker concentrations are to be avoided. At 65°C, more highly concentrated sulphuric acid attacks the amyl alcohol, producing olefins through dehydration which influence the result. Weaker concentrations reduce the oxidization effect. Destruction of the fat globule coating is incomplete which can lead to the formation of lumps.

Amyl alcohol

For the determination of fat content according to Gerber's method
An isomer mixture of 2-methylbutane-1-ol and 3-methylbutane-1-ol

Requirements:	Hazard symbol:	Hazard rating:
Density bei 20°C (0,811 ± 0,003) g ml ⁻¹		Xn R 10-20
• Boiling boundary: 98 % (by volume) has to distil at between 128°C and 132°C at 1 bar. • The amyl alcohol must not contain any substances which could influence the result. • A substitute for amyl alcohol can be used provided that it will bring about the same test results as would be achieved with amyl alcohol.		S 24/25 VbF A II

Please note:

- The isomers of amyl alcohol have different boiling points: 2-methylbutane-1-ol at 128°C and 3-methylbutane-1-ol at 132°C.
- Of the 8 known isomers of amyl alcohol, only this mixture is suitable for the Gerber method.
- Contamination with other isomers of amyl alcohol, particularly with the tertiary amyl alcohol 2-methylbutane-2-ol, produces false results. The obtained fat content result is too high.

REQUIRED EQUIPMENT

1. **Calibrated butyrometer with suitable stopper in accordance with**
DIN 12836-A 4, DIN 12836-A 6, DIN 12836-A 8, DIN-12836-A 5
2. **Pipette** DIN 10283-p **for milk** or **pipette** DIN 12837-A for milk
3. **Pipette** DIN 12837-B or 10 ml **measuring tap** for **sulphuric acid**
4. **Pipette** DIN 12837-C **for 1 ml amyl alcohol**
5. **Centrifuge for determining fat content in milk**, heatable, with rpm counter.
When used under a full load, this centrifuge must be capable of producing a centrifugal force of (350 ± 50) g on the inside of the butyrometer stopper within 2 minutes at the most. With a rotation radius of e.g. $(26 \pm 0,5)$ cm up to the inside of the butyrometer stopper, which is the distance between the point of torque and the butyrometer stopper, this centrifugal acceleration is reached at a rotation speed of (1100 ± 80) min⁻¹.
6. **Tempering devices for butyrometers**, e.g. a $(65 \pm 2)^{\circ}\text{C}$ water bath
With a heated centrifuge, a centrifuge bushing can be used to attach the butyrometer into the water bath. The read-off temperature must be $(65 \pm 2)^{\circ}\text{C}$.

PREPARATION OF THE TEST SPECIMEN

The milk in the specimen bottle is heated up to 20°C and thoroughly mixed through gentle shaking. This is done to bring about an even distribution of fat and to prevent frothing and the formation of butter.

Milk fat is lighter than water and creams if it is left standing. A fat-rich layer accumulates on the surface. Stirring and careful shaking restore the original distribution.

If the layer of cream cannot be evenly distributed in this way, the milk should be slowly heated to $35\text{--}40^{\circ}\text{C}$ and gently swirled around until a homogenous fat distribution is achieved. The milk is then cooled to 20°C before being drawn into the pipette.

Foam breaks the fat globule coating. The milk may begin to turn into butter when stirred and uniform distribution of fat is no longer possible.

The fat liquefies at $35\text{--}40^{\circ}$ and the distribution process is speeded up.

After the temperature has been set, the milk can stand for 3 to 4 minutes to allow any pockets of air to disperse.

The volumeters are calibrated at 20°C . Any variations in temperature will influence the volume. Air pockets reduce the density and hence the mass of milk measured.

CONDUCTING A TEST = WORK PROCEDURES

The same milk specimen must be tested twice.



Fig. 1 Protective glasses and rubber gloves must be worn when handling sulphuric acid

1. Place 2 butyrometers into a clamp (butyrometer stand). With the aid of a measuring tap, introduce 10 ml of sulphuric acid into the butyrometer without wetting the neck of the butyrometer (see Fig. 1).

2. Carefully turn the bottle with the milk specimen upside down three or four times. Then immediately pipette 10.75 ml of milk into the butyrometer in such a way that the butyrometer neck does not get wet and the milk does not mix with the sulphuric acid. To do so, lean the tip of the pipette laterally as deeply as possible against the wall of the butyrometer so that the milk forms a layer on top of the sulphuric acid (see Fig. 2).

When the Gerber method was first introduced, 11.0 ml of milk were used. By reducing the quantity of milk 10.75 ml, the determined fat content more closely matches the results of the reference method. If the butyrometer neck is wetted with milk, residues may cling to it.

A clear dividing line between the acid and the milk, without a brownish-coloured edge, is the sign of good layering.

3. 1 ml of amyl alcohol is pipetted onto the milk, or introduced by means of a measuring tap.

Due to the low density of amyl alcohol, the two liquids do not mix.

4. The butyrometer is sealed with a stopper without mixing the liquids.

As a rule, the lower end of the stopper comes into contact with the liquid.

5. The butyrometer is placed in the butyrometer tube with the bulb pointing downwards. Shake the butyrometer vigorously until the two liquids are thoroughly mixed. Keep your thumb firmly pressed down on the butyrometer stopper while doing so. Turn the butyrometer up and down several times in order to enable the sulphuric acid remaining in the bulb to disperse (see Fig. 3).



Fig. 3

The butyrometer is shaken in the tube (protective goggles and rubber gloves must be worn)

When the liquids are mixed, a considerable amount of heat is given off. The gas built up in this way can cause the stopper to shoot out, or the butyrometer to break.

The butyrometer tube is used as a safety precaution. Instead of using a butyrometer tube, the butyrometer can be wrapped in a cloth.

Too lax shaking of the butyrometer or holding it unnecessarily in a slanted position inhibits quick mixing and therefore the rapid oxidation of the entire specimen, thus ruining the careful work done layering the liquid.



Fig. 4 Filling the centrifuge



Fig. 5 The butyrometers are brought to the exact reading temperature in a water bath

6. Immediately after the mixture has been shaken and turned upside down a few times, the butyrometers, still hot and with the stoppers pointing downwards, are placed in buckets inside the heated Gerber centrifuge (see Fig. 4), whereby the butyrometers must be placed exactly opposite each other.

Beforehand, the column of fat should be set at the height of the expected level of fat content by turning the stopper.

After setting the time on the centrifuge, the centrifuge is started. After reaching a centrifugal force of (350 ± 50) g, which is reached as a rule after 1 minute, the corresponding speed of (1100 ± 50) rpm should be maintained for 4 minutes.

The centrifuge must be fitted with an interlocking lid. After the time set for the centrifuge has been reached, the rotor brake is automatically applied.

7. The butyrometers are now removed from the centrifuge, taking care not to tilt them, and are placed with the stoppers facing downwards for 5 minutes in a water bath heated to 65°C (see Fig. 5).

It is important to maintain an exact temperature so as to obtain accurate results. Only a read-off at 65°C will ensure an exact result. If the temperature is too low, the volume of the column is reduced and a fat content reading that is too low will be indicated.



Fig. 6 Measured values can be reliably and accurately read off with the aid of a safety reading lamp

RESULT AND DEGREE OF ACCURACY

8. After removing the butyrometer from the water bath, it should be held vertically at a height where the meniscus of the column of fat is at eye level. With the help of the stopper, mark the demarcation line between the residual mixture and fat on a whole subdivision of the butyrometer scale and read off the height of the fat column at the lowest point of the meniscus. If the reading takes too long, the butyrometer must be placed back into the water bath (see Figs. 6 and 7).

If the meniscus of the fat column is not at eye level, a parallax error results.




Fig. 7




Fig. 7a: 4,0-% reading




Fig. 7b: 3,95-% reading

The result should be read off to half a scale point, i.e. to 0.05 %. It is not possible to obtain a more accurate result with whole milk butyrometers. If the meniscus touches the graduation mark, then the result is to be recorded as such (Fig. 7a).

If the meniscus intersects the graduation mark, then the lower value is taken (Fig. 7b).

The difference between the readings from the two butyrometers must not be greater than 0.10 %, i.e. the reproducibility amounts to 0.10 %.

When recording the result, you must add the note "fat content according to the Gerber method".

If the two specimens differ by 0.1 %, the mean value of both readings is taken.

Specimen 1: 4.20 % | Specimen 2: 4.30 % | Result: 4.25 % fat content

However, if the two readings are 4.20 % and 4.25 % fat content, the lower value of 4.20 % is recorded – on the principle that it is better to err on the side of caution.

- After the first centrifuging, the two values were read off at 3.55 % and 3.60 %.
- After the second centrifuging, the two values were 3.60 % and 3.65 %. The homogenised milk fat content is stated to be 3.65 %.
- Should the difference still be greater than 0.05 % after the two last repetitions, i.e. after the 3rd and 4th centrifuging, the results of this test are to be discarded.

BUTYROMETRIC DETERMINATION OF THE FAT CONTENT OF VARIOUS DAIRY PRODUCTS

FOREWORD:

The butyrometric determination of the fat content in milk has been and is being replaced at a progressive rate by other routine tests (with appliances such as LactoStar, for instance). However, milk products such as cheese, ice cream, etc. either cannot be tested with such appliances or can only be tested after elaborate specimen preparation. For these products, butyrometric methods are a good alternative for routine analysis.

1.0 FIELD OF APPLICATION

Determination of fat content in milk and various dairy products.

2.0 VOLUMES

Unless stated otherwise, the following quantities of chemicals and test specimens are used:

Sulphuric acid:	10.0 ml (20°C + 2°C)
Amyl alcohol:	1.0 ml (20°C + 2°C)
Milk or dairy product:	10.75 ml (20°C + 2°C)

3.0 BRIEF DESCRIPTION OF THE BUTYROMETRIC DETERMINATION OF FAT CONTENT:

3.1 ... IN MILK (ACCORDING TO GERBER'S METHOD):

Perfectly cleaned milk butyrometers, most importantly free of fat residues, are filled in the following order: 10 ml of sulphuric acid (density: 1.818 +/- 0.003 g/ml), 10.75 ml of milk, and 1 ml of amyl alcohol. The milk and amyl alcohol are to be filled in layers so that they do not mix before shaking. After sealing the butyrometer, the contents are thoroughly mixed by shaking and turning the butyrometer upside down several times. Through careful adjustment of the sealing stopper, the contents of the butyrometer are regulated in such a way that the scale is full but no liquid enters the bulb. The butyrometer is centrifuged in the heated centrifuge, and then tempered in a 65°C water bath for 5 minutes. The dividing line between the sulphuric acid mixture and the fat column is set at a full scale line and the upper end of the fat column is read off at the lower meniscus.

3.2 ... IN HOMOGENISED MILK

As above, except that the specimens are centrifuged three times, for 5 minutes each time. In between centrifugings, the butyrometers are heated in a 65°C water bath for 5 minutes (for more detail, see page 19).

3.3 ... IN SKIM MILK AND WHEY

Use of skim milk butyrometers with narrowed scales according to Sichler's method.

Butyrometers are centrifuged twice and placed in a 65°C water bath for 5 minutes in between centrifugings.

3.4 ... IN CONDENSED MILK (UNSWEETENED)

The condensed milk is heated to 50°C and allowed to cool again. Then it is mixed with water in a 1:1 ratio. This dilution is then tested like milk according to Gerber's method. Fat content = read-off value x 2.

3.5 ... IN BUTTERMILK (MOHR AND BAUR'S MODIFICATION)

The butyrometer is filled with 10 ml of sulphuric acid (density: 1.830 +/- 0.003 g/ml). Instead of 10.75 ml, 10 ml of buttermilk and 2.0 ml of amyl alcohol are pipetted in. After sealing, the butyrometer is shaken and immediately centrifuged for 10 minutes. This prevents annoying blockages. The reading is taken after tempering the specimen at 65°C +/- 2°C. Fat content = read-off value x 1.075.

3.6 ... IN POWDERED MILK ACCORDING TO TEICHERT'S METHOD

Use of powdered milk butyrometer according to Teichert's method.

The butyrometer is filled with 10 ml of sulphuric acid (density: 1.818 +/- 0.003 g/ml). 7.5 ml of water and 1 ml of amyl alcohol are layered onto it. 2.5 g of powdered milk are weighed in a weighing boat and transferred into the butyrometer through a funnel, using a hair brush. After sealing, the butyrometer is thoroughly shaken, and intermittently placed in a 65°C water bath several times. It is centrifuged twice, 5 minutes each time, in the heated centrifuge and the value is read off after the butyrometer is placed in the water bath for 5 minutes.

3.7 ... IN CREAM ACC. TO ROEDER'S METH. (WEIGHING METH.)

Use of cream butyrometer according to Roeder's method.

5 g of cream are weighed out into the glass beaker situated in the stopper and transferred into the butyrometer. Sulphuric acid (density: 1.522 +/- 0.005 g/ml) is poured through the upper opening of the butyrometer until it reaches just over the upper edge of the glass beaker. After sealing, the butyrometer is placed in a 70°C water bath and shaken repeatedly until the protein is completely dissolved. Sulphuric acid is poured in until it reaches the beginning of the scale, 1 ml of amyl alcohol is added and the butyrometer is sealed, shaken, and placed in the 70°C water bath for another 5 minutes. Then, it is centrifuged for 5 minutes and tempered in a 65°C water bath. The reading is taken at 65°C, the fat column adjusted to the zero point, the value read off at the lower meniscus.

3.8 ... IN CREAM ACC. TO SCHULZ-KLEY'S METH. (WEIGHING METH.)

Use of cream butyrometer according to Schulz-Kley's method.

10 ml of sulphuric acid (density: 1.818 +/- 0.003 g/ml), 5 ml of water, ca. 5 g of cream measured by differential weighing using a syringe or weighing pipette attached to the scale, and 1 ml of amyl alcohol are successively introduced into the butyrometer. After sealing, the contents are mixed by shaking and turning the butyrometer upside down and centrifuged for 5 minutes in the heated centrifuge. The value is read off after 5 minutes of tempering in a 65°C water bath. Note: do not allow more than 15 minutes to pass between adding the water and shaking the specimen due to the possibility of a decrease in the heat of reaction as a result of the addition of water. The dissolution process must be completed within 60 seconds.

Fat content = read-off value x 5 / cream net weight.

3.9 ... IN CREAM ACC. TO KOEHLER'S METH. (MEASURING METH.)

Use of cream butyrometer according to Koehler's method.

10 ml of sulphuric acid (density: 1.818 +/- 0.003 g/ml), 5.05 ml of cream, 5 ml of water, and 1 ml of amyl alcohol are poured in turn into the cream butyrometer. When using a cream syringe, it must be rinsed several times with water before the 5 ml of water are introduced. The butyrometer is sealed, shaken, centrifuged for 5 minutes, and then the value is read off after tempering the specimen in a 65°C water bath. The value is read off from the zero point.

3.10 ... IN CHEESE ACC. TO VAN GULIK'S METH. (WEIGHING METH.)

[see ISO 3433] Use of the cheese butyrometer according to Van Gulik's method.

After pouring 15 ml of sulphuric acid (density: 1.522 +/- 0.005 g/ml) into the Van Gulik butyrometer, closed at the scale end, 3 g (+/- 0.2 g) of cheese are added by means of a weighing boat and hair brush, and then the filler opening is sealed. Pasty cheese samples must be weighed out into the perforated glass beaker which accompanies the Van Gulik butyrometer and introduced into the butyrometer. The sealed butyrometer is placed in a 70°C - 80°C water bath with the scale pointing upwards and shaken repeatedly until the cheese is dissolved. Afterwards, 1 ml of amyl alcohol is added, followed by sulphuric acid until it approximately reaches the 15 % mark of the scale. Then close the butyrometer, mix the contents, temper for 5 minutes in a 65°C water bath, adjust the fat column to the zero point and read off the absolute fat content. The reading is taken from the lower end of the meniscus.

3.11 ... IN ICE CREAM ACC. TO KOEHLER'S METH. (MEASURING METH.)

Use of the ice cream butyrometer according to Koehler's method.

Any icing or rough particles (e.g. fruit, etc.) must be removed. The ice cream is thoroughly mixed after it has reached room temperature; possible air pockets can be almost completely removed by evacuation.

10 ml of sulphuric acid (density: 1.818 +/- 0.003 g/ml), 5 ml of ice cream, 5 ml of water, and 1 ml of amyl alcohol are introduced in turn into the ice cream butyrometer. When using a syringe, it must be rinsed several times with water before the 5 ml of water are introduced. Should the butyrometer prove to not be sufficiently full, up to 2 ml of water can be added. The butyrometer is sealed, shaken, centrifuged for 5 minutes, and, after tempering in a 65°C water bath for 5 minutes, the value is taken.

3.12 ... IN ICE CREAM ACC. TO ROEDER'S METH. (WEIGHING METH.)

Use of ice cream butyrometer according to Roeder's method.

5 g of well-mixed ice cream are weighed out into the glass beaker situated in the stopper and transferred into the butyrometer. Sulphuric acid (density: 1.522 +/- 0.005 g/ml) is poured through the upper opening of the butyrometer until it reaches just over the upper edge of the glass beaker. After sealing, the butyrometer is placed in a 70°C water bath and shaken repeatedly until the protein is completely dissolved. 1 ml of amyl alcohol is added, and then sulphuric acid is poured in until it reaches the 10 % mark on the scale. The butyrometer is sealed, shaken, and placed in the 70°C water bath for an additional 10 minutes. During this time, the butyrometer is shaken at regular intervals. This process is followed by 7 minutes of centrifuging and tempering in a 65°C water bath. The value is taken at 65°C, the fat column adjusted to the zero point. The value is read off at the lower meniscus.

3.13 ... IN BUTTER ACC. TO ROEDER'S METH. (WEIGHING METH.)

Use of butter butyrometer according to Roeder's method.

5 g of butter are weighed out into the glass beaker situated in the stopper and transferred into the butyrometer. Sulphuric acid (density: 1.522 +/- 0.005 g/ml) is poured through the upper opening of the butyrometer until it reaches just over the upper edge of the glass beaker. After sealing, the butyrometer is placed in a 70°C water bath and shaken repeatedly until the protein is completely dissolved. Sulphuric acid is poured in until it reaches the beginning of the scale, followed by 1 ml of amyl alcohol. The butyrometer is sealed, shaken, and placed in the 70°C water bath for an additional 10 minutes. This process is followed by 5 minutes of centrifuging and approximately 5 minutes of tempering in a 65°C water bath. The value is read off at 65°C at the lower meniscus.

3.14 ... IN MAYONNAISE ACC. TO ROEDER'S METH. (WEIGHING METH.)

Use of butter butyrometer according to Roeder's method.

1 g of mayonnaise is weighed out into glass beaker situated in the stopper and transferred into the butyrometer. Sulphuric acid (density: 1.522 +/- 0.005 g/ml) is poured through the upper opening of the butyrometer until it reaches just over the upper edge of the glass beaker. After sealing, the butyrometer is placed in a 70°C water bath for 30 minutes and shaken repeatedly until the protein is completely dissolved. Sulphuric acid is poured in until it reaches the beginning of the scale, followed by 1 ml of amyl alcohol. The butyrometer is sealed, shaken, and placed in the 70°C water bath for an additional 5 minutes. This process is followed by 10 minutes of centrifuging and approximately 5 minutes of tempering in a 65°C water bath. The value is read off at 65°C at the lower meniscus. The read-off value is multiplied by 5 to obtain the correct fat content.

3.15 BUTYROMETRIC DETERMINATION OF THE FAT CONTENT OF MEAT AND SAUSAGE ACC. TO GERBER'S METHOD (VAN GULIK)

According to the recommended methodology of "Pohja and Associates".

Instruments:

1. Butyrometer

Cheese butyrometer according to Van Gulik's method

2. Centrifuge

Milk centrifuge with an RCF (relative centrifugal force) of 350 g +/- 50 g
(e.g. SuperVario-N or Nova Safety)

3. Water bath

Shaking water bath with a temperature of 65°C +/- 2°C

4. Precision scale

5. Appliances for sampling preparation

A mixer or the like is recommended for reducing or homogenising the sample.

Chemicals:

1. Sulphuric acid

Density of (1,818 +/- 0,003) g ml⁻¹ at 20°C, colourless or only slightly discoloured and free from any substances which might influence the outcome

2. Amyl alcohol

Density of (0,811 +/- 0,003) g ml⁻¹ at 20°C

Procedure:

Firstly, the perforated glass cylinder (cheese beaker) of the cheese butyrometer is fitted into the butyrometer stopper (this stopper is fashioned with a hole). Then, exactly 2,500 g of the homogenised sample is weighed and introduced into the glass cylinder ("cheese beaker"). The cheese beaker and stopper are attached to the body of the butyrometer. 10 ml of sulphuric acid, diluted with water in a volume ratio of 1:1, is poured into the small opening on the upper end of the butyrometer. The small opening is sealed with the small compatible stopper and placed in a shaking water bath, where it is left at 65°C for approximately 30 to 40 minutes, until all proteins are dissolved. Then, 1 ml of amyl alcohol is pipetted into the butyrometer and, after renewed sealing, is shaken powerfully. Afterwards, sulphuric acid is added until the overall fluid level is at about 30% on the scale. Then the butyrometer is centrifuged for 5 minutes at 350 g. Note: the centrifuge must be loaded evenly. I.e. one butyrometer cannot be centrifuged

alone – this leads to an unbalance and the risk of glass breaking. The butyrometers are then placed in the water bath to be tempered, i.e. the shaking mechanism is turned off. The reading should be taken immediately after the specimens are taken out of the water bath as the fat column diminishes significantly at very small temperature decreases, resulting in too low fat content read-off values. The butyrometers are dimensioned for samples of 3,000 g, meaning that the determined values are to be assessed at this substance amount (increase by 16.666 %).

BUTYROMETER



The foundation of the Gerber method is the butyrometer. The original butyrometer with the rounded neck, invented by Dr. N. Gerber, was developed into the well-known flat butyrometer under the management of Paul Funke, accompanied by his glassblowers. While the original Gerber butyrometer is hardly employed anymore, the **original FUNKE-GERBER** butyrometers with the flattened scale neck are used almost exclusively. The flattened scale neck increases comfort when reading off values and improves precision.

These flat butyrometers are manufactured to standards of unmatched quality and the highest production control. Each individual butyrometer is individually gauged and correspondingly scaled. The high level of precision in setting the scale divisions and volumes guarantees exact test results.

FUNKE-GERBER butyrometers are precision instruments with a flattened scale section, manufactured from acid-proof glass (borosilicate) in compliance with national (DIN) and international (ISO/IDF etc.) standards. Our over 100 years of production experience and high assembly numbers enable us to offer the highest quality at low prices. You will find a multitude of different butyrometers for various tasks in the following pages of this catalogue.



In Germany and some other countries, butyrometers must be officially calibrated. These butyrometers are labelled with an engraved mark (see adjacent figure). Although all other butyrometers are not officially calibrated, they are manufactured in the same way and meet the same high quality standards.

All butyrometers come in standard packs of 10.
Please place your order in units of 10.

Precision butyrometer

for drinking milk and vat milk, frosted rear scale wall,
fault tolerance 0.025%

3150	0 – 4 %: 0.05 (accessory: 3280)
-------------	---------------------------------

Milk butyrometer

3151	0 – 5 %: 0.1 (accessory: 3280)
-------------	--------------------------------

3152	0 – 6 %: 0.1 (accessory: 3280)
-------------	--------------------------------

3153	0 – 7 %: 0.1 (accessory: 3280)
-------------	--------------------------------

3154	0 – 8 %: 0.1 (accessory: 3280)
-------------	--------------------------------

3155	0 – 9 %: 0.1 (accessory: 3280)
-------------	--------------------------------

3156	0 – 10 %: 0.1 (accessory: 3280)
-------------	---------------------------------

3157	0 – 12 %: 0.1 (accessory: 3280)
-------------	---------------------------------

3158	0 – 16 %: 0.2 (accessory: 3280)
-------------	---------------------------------



Skim milk butyrometer

according to Sichler's method, with rounded scale

3160	0 – 1 %: 0.01, with open bulb (accessories: 3280, 3290)
-------------	--

3160-G	0 – 1 %: 0.01, with closed bulb (accessory: 3280)
---------------	--



Skim milk butyrometer

according to Kehe's method

3161	0 – 4 %: 0.05 (accessory: 3280)
-------------	---------------------------------

3162	0 – 5 %: 0.05 (accessory: 3280)
-------------	---------------------------------

Skim milk butyrometer

according to Siegfeld's method

3164 0 – 0.5 %: 0.02 (accessory: 3280)



Powdered milk butyrometer

according to Teichert's method

3170 0 – 35 %: 0.5, (accessory: 3310)

3171 0 – 70 %: 1.0, (accessory: 3310)



Ice cream and condensed milk butyrometer

according to Roeder's weighing method

3180 0 – 6 – 12 %: 0.1, (accessories: 3290, 3300, 3320)

3181 0 – 15 %: 0.2, (accessories: 3290, 3300, 3320)



Cream butyrometer

measuring method, for ice cream

3189 0 – 15 %: 0.2 (accessory: 3280)

3190 0 – 20 %: 0.2 (accessory: 3280)



Cream butyrometer

according to Roeder's weighing method



3200	0 – 5 – 40 %: 0.5 (accessories: 3290, 3300, 3320)
3201	0 – 30 – 55 %: 0.5 (accessories: 3290, 3300, 3320)
3202	0 – 50 – 75 %: 0.5 (accessories: 3290, 3300, 3320)
3203	0 – 5 – 70 %: 1.0 (accessories: 3290, 3300, 3320)

Cream butyrometer

according to Schulz-Kley's weighing method
with closed bulb



3208	0 – 5 – 40 %: 0.5 (accessory: 3280)
-------------	-------------------------------------

Cream butyrometer

according to Koehler's measuring method



3209	0 – 30 %: 0.5 (accessory: 3280)
3210	0 – 40 %: 0.5 (accessory: 3280)
3211	0 – 50 %: 1.0 (accessory: 3280)
3212	0 – 60 %: 1.0 (accessory: 3280)
3213	0 – 70 %: 1.0 (accessory: 3280)
3214	0 – 80 %: 1.0 (accessory: 3280)

Butter butyrometer

according to Roeder's weighing method

3220 0 – 70 – 90 %: 0.5 (accessories: 3290, 3300, 3323)

Cheese butyrometer

according to Van Gulik's weighing method

3230 0 – 40 %: 0.5 (accessories: 3290, 3300, 3321)



Curd butyrometer

weighing method

3240 0 – 20 %: 0.2 (accessories: 3290, 3300, 3321)

Food butyrometer

according to Roeder's weighing method

3250 0 – 100 %: 1.0 (accessories: 3290, 3300, 3320)

Free fat butyrometer

for determining free fat content in milk and cream,

3252 complete with screw cap, scale 0.002 g

Babcock bottle

without stopper

3254 0 – 8 % for milk, stopper on request**Babcock bottle**



without stopper

3256 0 – 20 % for cream (accessory: 3290)**Babcock bottle**


without stopper

3258 0 – 60 % for cream and cheese (accessory: 3290)**Patent closure FIBU**

for all measuring method butyrometers

3260 FIBU without adjustment key
(Fig. with adjustment key art. no. 3270)**Patent closure GERBAL****3261** for all measuring method butyrometers**Patent closure NOVO****3262** for all measuring method butyrometers

3270 **Adjustment key**
for patent closure FIBU



3271 **Adjustment key**
for patent closure GERBAL

3272 **Adjustment key**
for patent closure NOVO

3280 **Rubber stopper, conical**
for all measuring method butyrometers
11 x 16 x 43 mm



3290 **Rubber stopper** for sealing the bulbs
of all weighing method butyrometers
9 x 13 x 20 mm



3300 **Rubber stopper with hole**
for all weighing method butyrometers
17 x 22 x 30 mm



3310 **Rubber stopper without hole**
for powdered milk butyrometer
(also suitable for the extraction tube acc. to Mojonier
art. no. 3870, 3871)
17 x 22 x 30 mm



3315 **Glass nail**
for powdered milk butyrometer
length: 41.5 mm




Cream beaker, unperforated

for ice cream and condensed milk
butyrometers and cream butyrometers
according to Roeder's method

3320

$l_3 = 75 \text{ mm}$, $l_2 = 49 \text{ mm}$, $l_1 = 26 \text{ mm}$, $d_2 = 15 \text{ mm}$, $d_1 = 5 \text{ mm}$



Cheese beaker, perforated

for butyrometers according to Van Gulik's method

3321

$l_3 = 75 \text{ mm}$, $l_2 = 49 \text{ mm}$, $l_1 = 26 \text{ mm}$, $d_2 = 15 \text{ mm}$, $d_1 = 5 \text{ mm}$

Cheese beaker, perforated, short design

for butyrometers according to Van Gulik's method

3321-001

$l_3 = 66 \text{ mm}$, $l_2 = 38 \text{ mm}$, $l_1 = 27.8 \text{ mm}$, $d_2 = 15 \text{ mm}$, $d_1 = 5 \text{ mm}$



Weighing boat for butter

for butyrometers according to Roeder's method

3322

$l_3 = 75 \text{ mm}$, $l_2 = 45 \text{ mm}$, $l_1 = 30 \text{ mm}$, $d_2 = 15 \text{ mm}$, $d_1 = 5 \text{ mm}$



Butter beaker with 2 holes

$l_3 = 75 \text{ mm}$, $l_2 = 48 \text{ mm}$, $l_1 = 27 \text{ mm}$, $d_2 = 15 \text{ mm}$, $d_1 = 5 \text{ mm}$

3323




Cleaning brush
for butyrometer body

3324 length: 270 mm



Cleaning brush
for butyrometer neck


3325 length: 278 mm



Butyrometer stand
(also suitable for special solubility index tubes, art. no. 3637)

3330 for 36 samples (PP plastic)

3331 for 12 samples (PP plastic)



Shaking stand
for 12 samples (PP plastic)


3332

Protective shaking hood
for 36 samples (PP plastic),
compatible with art. no. 3330

3340

for 12 samples (PP plastic)
compatible with art. no. 3331

3341



Permanent automatic dispenser
with ground-in measuring chamber and stopper,
one spout in accordance with DIN 10282

3390 10 ml sulphuric acid

3391 1 ml amyl alcohol

Stand for permanent automatic dispenser
consists of stand panel, stem and retaining ring
with socket

3400 10 ml for 1 permanent automatic dispenser

3401 1 ml for 1 permanent automatic dispenser

3402 10 ml / 1 ml for 2 permanent automatic dispensers



Automatic tilt measure Superior

with rubber stopper and storage bottle 500 ml / 250 ml

3420	10 ml sulphuric acid
-------------	----------------------

3421	1 ml amyl alcohol
-------------	-------------------



Weighing pipettes

curved

3425	1 ml , d = 6 mm
-------------	------------------------

3426	2 ml , d = 8 mm
-------------	------------------------

3427	3 ml , d = 9 mm
-------------	------------------------

3428	5 ml , d = 6 mm
-------------	------------------------

3429	10 ml , d = 7 mm
-------------	-------------------------

Volumetric pipettes

with one ring mark

3430	10 ml sulphuric acid
-------------	----------------------

3431	10.75 ml milk
-------------	---------------

3432	11 ml milk
-------------	------------

3433	1 ml amyl alcohol
-------------	-------------------

3434	5.05 ml cream
-------------	---------------

3435	5 ml water
-------------	------------

3436	5 ml cream
-------------	------------

3437	50 ml, short design
-------------	---------------------

3438	25 ml, short design
-------------	---------------------



Syringes

nickel-plated brass



3440	10.75 ml milk
3441	10.75 ml milk, rep. exch.
3442	5.05 ml cream
3443	5.05 ml cream, rep. exch.
3450	11 ml milk
3452	5 ml cream

Pipette stand

3460 PVC, for pipettes of various sizes



Cleaning brush

for pipettes

3470 length: 470 mm

3480 Safety goggles

LactoStar

(article no. 3510)



THE NEW GENERATION OF INSTRUMENTS

Chemical milk analysis device with fully automatic cleaning and rinsing system and zero point calibration for the fast and accurate testing of milk

Countless installations in institutes and laboratories all over the world attest to the outstanding quality, reliability and accuracy of this chemical analysis device.

The following parameters can be determined quickly and reliably with just one measurement:

Contents	Measuring range	Reproducibility (r)
Fat:	0.00 % to 40.00 %	± 0.02 %*
Protein:	0.00 % to 10.00 %	± 0.03 %
Lactose:	0.00 % to 10.00 %	± 0.03 %
SNF:(fat free dry matter)	0.00 % to 15.00 %	± 0.04 %
Minerals/conductance	0.01 % to 5.00 %	± 0.02 %

* The reproducibility equals +0.02 % in the 0% to 8 % fat range.
In the higher ranges of 8 % to 40 % fat, the reproducibility equals 0.2 %

The measurement resolution is 0.01 %.

The accuracy depends on the respective calibration.


Further parameters are determined based on calculational algorithms:

- Density (calculated value)
- Freezing point (calculated value)

The software is continually improved with the aim of obtaining additional interesting parameters. The updates are quickly and easily transmitted over the interfaces. In this way, the instrument is kept up-to-date over a long period of time.

High matrix tolerance

The device is characterised by a high matrix tolerance which it owes to the multi-sensor measurements systems used. This means different kinds of milk can be measured with the same calibration (product profile).



Operation

The operation is clear and simple.

5-key menu-driven handling: 4 arrow keys and one "enter" key.

By pressing the "enter" key, the function or action, which has been selected with the help of the arrow keys, is started.

LactoStar




Language selection

You can select your preferred language from a number of menu languages. Currently, German, English, French, Spanish, among others, are available. The number of languages is continuously being expanded with help from our partners in different countries. The language is selected in the same way as with the previous settings.



Calibration

For customer-specific calibration, the already existing basic calibrations are merely adjusted. This is done with a simple two point calibration (A calibration and B calibration). Each parameter is calibrated in only one step. A clear calibration menu simplifies the entry of reference values. Up to 20 different calibration data sets can be saved. Therefore, you can switch from one product to another (e.g. from milk to cream) without having to re-calibrate.



Automatic maintenance


The instrument has three pumps, namely the measuring pump, the rinsing pump, and the cleaning pump, all of which are connected to the corresponding canisters.

Up to five different times can be entered for various maintenance activities: rinsing, cleaning, and zero point calibration. Thus routine tasks are completed automatically.


Replacing the pump heads

The pumps are located under the stainless steel cover sheet on the left side of the instrument. The pump heads can be replaced easily without the use of tools.

The old pump heads can be pulled off by pushing in on both tabs (see Fig, "X" and "Y") simultaneously. The new pump heads are set on the motor axis and pressed down on until both tabs snap into place.



Principle design



Technical specifications:

Sample capacity:	up to 90/h
Sample volume:	from 12 ml to 20 ml
Interfaces:	1 x parallel, 1 x serial (RS 232 / 9.600 baud), USB 6 volt electrical power supply for thermal printer (order no. 7151)
Connection values:	230V / 115V AC (50...60 Hz) 180 W
Dimensions:	43 x 20 x 43 cm (B x H x T)
Weight:	approx. 15.7 kg (net)

Ordering data

Article no.	Description
3510	LactoStar
7151 *	Thermal printer, incl. 1 thermal paper roll
3511 *	Each canister 5 L, for rinsing water and cleaning agent
3516 *	Hardware standardisation, 250 ml
3563 *	Cleaning agent, 500 ml

(articles marked with * are included in the 3510 scope of delivery)

Accessories (optional)

3040	Milk sample bottle without metal bottom, 80 ml / PE
3041	Milk sample bottle with metal bottom, 50 ml / PP
7157	Thermal paper roll for thermal printer

Spare and wear parts

3510-023	Hose pump, complete
3510-023 A	Pump head (attachment for hose pump)

LactoFlash

(article no. 3530)



Inexpensive chemical analysis device for the fast and accurate determination of fat and SNF content

Countless installations in institutes and laboratories all over the world attest to the outstanding quality, reliability and accuracy of this chemical analysis device.

The following parameters can be determined quickly and reliably with just one measurement:

Parameter	Resolution	Reproducibility (r)	Measuring range
Fat:	0.01 %	0.02 % in 0 ... 5 % range 0.2 % in 5 ... 30 % range	0 ... 30 %
SNF:	0.01 %	0.04 %	0 ... 15 %

Further parameters are determined based on calculational algorithms:

Parameter	Resolution	Reproducibility (r)	Measuring range
Density:	0.0001	0.001	no limit
Protein:	0.01 %	0.03 %	no limit / calculated value
Lactose:	0.01 %	0.02 %	no limit / calculated value
Gpp:	0.001°C	0.002°C	no limit / calculated value

Quick and easy replacement of pump heads and measuring cells.

The pump head (wear part) can be replaced easily without the use of tools. This is done by removing the blue side cover sheet and pulling off the old pump head by pushing in on both side tabs simultaneously. The new pump head is set on and pressed down on until both tabs snap into place.



If one of the two measuring cells has to be replaced, it can simply be pulled out of the plug connection. The new measuring cell is plugged in.

LactoFlash

Operation



The instrument has 4 arrow keys and an "enter" key. With the "enter" key, the function or action, which has been selected with the help of the arrow keys, is started.



Language selection

Two menu languages are available: German and English.



Calibration

For customer-specific calibration, the already existing basic calibrations are merely adjusted. This is done with a simple two point calibration (A calibration and B calibration). Each parameter is calibrated in only one step. A clear calibration menu simplifies the entry of reference values.

Technical specifications:

Sample capacity:	up to 120/h
Sample volume:	from 12 ml to 20 ml
Interfaces:	1 x parallel, 1 x serial (RS 232 / 9.600 baud) 6 volt electrical power supply for thermal printer (order no. 7151)
Connection values:	230V / 115V AC (50..60 Hz) 60 W
Dimensions:	30 x 24 x 33 cm (w x h x d)
Weight:	5 kg (net)

Ordering data

Article no.	Description
3530	LactoFlash
7151	Thermal printer, incl. 1 thermal paper roll
3516	Hardware standardisation
3563	Cleaning agent, 500 ml

Accessories (optional)

3040	Milk sample bottle without metal bottom, 80 ml / PE
3041	Milk sample bottle with metal bottom, 50 ml / PP
7157	Thermal paper roll for thermal printer

Spare and wear parts

3530-023	Hose pump, complete
3530-023 A	Pump head (attachment for hose pump)

LactoStar

newly developed instrument
for the routine testing of milk
fat, protein, lactose, SNF, freezing point

see p. 36 for a more detailed description
accessories included

3510

Accessories:

Thermal printer	art. no. 7151
Canister	art. no. 3511
Hardware standardisation	art. no. 3516
Cleaning agent	art. no. 3563

Replacement parts:

3510-023	Hose pump, complete
3510-023A	Attachment for hose pump



Canister

space saving design, 5 L

3511

334 x 64 x 334 mm (w x d x h)



Hardware standardisation,

for art. no. 3510, 3530

3516

250 ml



REFERENCE MATERIAL

Reference milk, 1.5 % fat class

The exact values depend on the batch.

3517

They are included in the delivery.

Reference milk, 3.5 % fat class

The exact values depend on the batch.

3518

They are included in the delivery.

Reference cream, 30 % fat class

The exact values depend on the batch.

3519

They are included in the delivery.

Reference milk, 0.1 % fat class

The exact values depend on the batch.

3521

They are included in the delivery.

LactoFlash

Chemical analysis device for the fast and accurate determination of fat and SNF content. accessories included

3530

Accessories:

Thermal printer:	art. no. 7151
Hardware standardisation	art. no. 3516
Cleaning agent	art. no. 3563

Replacement parts:

3530-023	Hose pump, complete
3530-023A	Attachment for hose pump



Shaking water bath

stainless steel with cover,
shaking stand and 18 tubes

Technical specifications:

PID controller with PT-100 temperature sensor	
Setting:	in 0.1°C increments
Accuracy:	+/- 0.1°C
Connection values:	230 V / 8.7 A, 2000 W
Volume:	22 L
Internal dimensions:	350 x 290 x 220 mm
External dimensions:	578 x 436 x 296 mm
Weight:	approx. 17 kg net

3550

Butyrometer bucket

pressure cast light metal
accessory for SuperVario-N centrifuge
(art. no. 3680 p. 48)

3631 1 bucket

3631-12 set with 12 buckets

3631-24 set with 24 buckets

3631-36 set with 36 buckets

**Babcock bucket**

accessory for SuperVario-N centrifuge
(art. no. 3680)

3632

**Bucket for ADPI tubes**

accessory for SuperVario-N centrifuge (art. no. 3680)

3633

**Solubility index tube**

ADPI, 50 ml, graduated from 0 – 20 ml,
mark at 50 ml
see SuperVario-N (art. no. 3680)

3634

**Stand**

for 6 tubes (art. no. 3634)

3636

Special solubility index tubes

for determining the solubility of powdered milk
fit in butyrometer tube (art. no. 3641)
for use in the Nova Safety bench centrifuge
(art. no. 3670)

3637

for compatible rubber stoppers, see art. no. 3050
for compatible stand, see art. no. 3331



3638 **Centrifuge tube**
with 2 stoppers according to Friesse's method



3639 **Homogenisation pipette**
with a mark at 5 ml and 25 ml, incl. stopper

external diameter: 24 mm
length without stopper: 152 mm



3641 **Replacement butyrometer tube**
for Nova Safety art. no. 3670

brass, with flanged edge
external diameter: 27 mm
internal diameter: 25.8 mm
length: 170 mm



Nova Safety centrifuge

Thousands of these centrifuges have been installed in laboratories all over the world. They are characterized by extreme robustness and reliability. This table centrifuge with the angular rotor can be used for the determination of fat content according to Dr. N. Gerber's method as well as for the determination of the solubility of powdered milk.

Properties:

Automatic interlocking lid
Automatic brake (braking time < 8 s)
Centrifugation timer (digital)
Heater, thermostatic set point at 65°C
Capacity: max. 8 butyrometer

Technical specifications:

RCF: 350 g +/- 50 g
rpm: 350 U/min
Effective radius: 160 mm
Weight: 13 kg
Dimensions (l x w x h): 470 x 380 x 230 mm

3670



SuperVario-N



MULTIPURPOSE CENTRIFUGE FOR THE DAIRY LABORATORY

This centrifuge stands out due to its exceptional engine smoothness. That it is largely free of vibration and has swing out butyrometer buckets positively effects the life-time of your butyrometers. Correspondingly, good results (reproducibility and comparability) are assured. For these reasons, the SuperVario-N is often used as a reference centrifuge for calibration purposes. Due to its versatility, the SuperVario-N is widely accepted in dairy laboratories. High versatility means free programmability of rpm, temperature and time ("free mode") as well as 4 set programs for the following tests:

Dr. N. Gerber's method (determination of fat content)
 Roese-Gottlieb's method (determination of fat content, reference method)*
 Babcock's method (determination of fat content)
 Solubility (determination of the solubility of powdered milk)

* Operation only possible under observation of respective safety regulations

Properties:

- Stainless steel housing
- Programmable rotor speed from 600 to 1130 rpm
in increments of 10 rpm
(corresponds to a g-value of 77 to 372 g)
- Programmable heater up to 68°C in 1°C increments
- Programmable centrifugation time from 1 to 99 minutes
- Automatic interlocking safety lid
- Automatic shut down in case of unbalance
- Automatic brake

Technical specifications:

Connection values:	230 V/50 ... 60 Hz/1200 VA
Weight, empty:	26 kg
Total height with lid:	460 mm
Filling height:	370 mm
Rotor speed range:	600 to 1130 rpm**
Temperature range:	room temperature up to 68°C

** For the determination of fat content due to Gerber's method, a g-value of 350 g ± 50 g is required. With a relative centrifugal force (RCF) of 365 g in its unloaded state (idle running) and 340 g in its fully loaded state, the SuperVario-N fulfils the standard specifications in an exemplary manner.

MILK LABORATORY CENTRIFUGES

Centrifuges for butyrometric determination of fat content according to Dr. N. Gerber
K. Schaefer, graduate engineer, reports

QUIET OPERATION

In order to avoid glass breakage and to increase butyrometer lifetime, it is very important that the centrifuge operates with the lowest level of vibration possible. The different types of centrifuges are:

TYPE 1: Centrifuge with flat-lying butyrometers

This way of loading butyrometers guarantees that they will be gently treated during centrifugation. However, this type of centrifuge tends to lead to a renewed intermixing of the separated phases after the centrifugal process.

TYPE 2: Centrifuge with angular rotor

The butyrometers are held in the angular rotor at a fixed angle. Unfortunately, this causes the long, thin butyrometer necks considerable stress. This design is predominately found in small, inexpensive centrifuges.

TYPE 3: Centrifuge with swing-out butyrometer holders

The butyrometers swing out horizontally in mounted movable holders. The butyrometers are only stressed along their vertical axis. For this reason, this type of centrifuge is preferable.

These special centrifuges differ from other laboratory centrifuges in several ways. The following points should be taken into consideration when purchasing and using centrifuges for the determination of fat content according to Dr. N. Gerber's method:

UNBALANCE

The centrifuge should be equipped with an unbalance shut-down mechanism. In case of glass breakage (butyrometer breakage) or other types of unbalance, the centrifuge shuts itself off automatically.

INTERLOCKING LID

For safety purposes, more and more centrifuges are being equipped with an interlocking lid.

HEATING

Heating a centrifuge reduces butyrometer cooling. This means that the subsequent tempering time in the water bath can be kept to a minimum and leads to a more reliable realization of the analysis. The temperature in the centrifuge tank should be at least 50°C.

SET-UP

The centrifuge must be set up on a flat, secure surface (e.g. a stable tabletop or platform). The lowest possible humidity and a room temperature of less than 30°C are preferable.

ROUTINE OPERATION/MAINTENANCE

The centrifuge should be loaded in such a way that it is as balanced as possible, i.e. the butyrometers must always be evenly positioned. In case of glass breakage, the centrifuge should be cleaned immediately after standstill is reached. This prevents unnecessary corrosion and guarantees a long lifetime.

RPM

The determination of fat content according to Gerber's method specifies a "RCF" (relative centrifugal force) of 350 g with a maximum variation of ± 50 g. The RCF does not only depend on the rpm but also on the effective radius. The effective radius is defined as the distance between the rotor and the outer end of the butyrometer. For this reason, the rpm of different centrifuge types varies as a function of their respective radii. What is important is that the rpm is constant or only changes negligibly (within the range of tolerance, see above), depending on whether the centrifuge is fully or only partially loaded.

The RCF is calculated in the following way:

$$RCF = 1,12 \times 10^{-6} \times R \times N^2$$

$$N = \sqrt{\frac{RCF}{1,12 \times 10^{-6} \times R}}$$

whereby:

R = effective horizontal radius in mm;

N = revolutions per minute [min^{-1}].

SYNOPTICAL TABLE OF THE DEPENDENCE OF G-FORCE AND RPM

rpm (min^{-1})	Head A ($\phi=52$ cm) g force	Head B ($\phi=38$ cm) g force	Head C ($\phi=38$ cm) g force
600	104.8 g	76.6 g	76.6 g
610	108.4 g	79.2 g	79.2 g
620	111.9 g	81.8 g	81.8 g
630	115.6 g	84.5 g	84.5 g
640	119.3 g	87.2 g	87.2 g
650	123.0 g	89.9 g	89.9 g
660	126.8 g	92.7 g	92.7 g
670	130.7 g	95.5 g	95.5 g
680	134.7 g	98.4 g	98.4 g
690	138.6 g	101.3 g	101.3 g
700	142.7 g	104.3 g	104.3 g
710	146.8 g	107.3 g	107.3 g
720	151.0 g	110.3 g	110.3 g
730	155.2 g	113.4 g	113.4 g
740	159.5 g	116.5 g	116.5 g
750	163.8 g	119.7 g	119.7 g
760	168.2 g	122.9 g	122.9 g
770	172.7 g	126.2 g	126.2 g
780	177.2 g	129.5 g	129.5 g
790	181.7 g	132.8 g	132.8 g
800	186.4 g	136.2 g	136.2 g
810	191.1 g	139.6 g	139.6 g
820	195.8 g	143.1 g	143.1 g
830	200.6 g	146.6 g	146.6 g
840	205.5 g	150.2 g	150.2 g
850	210.4 g	153.7 g	153.7 g
860	215.4 g	157.4 g	157.4 g
870	220.4 g	161.1 g	161.1 g
880	225.5 g	164.8 g	164.8 g
890	230.7 g	168.6 g	168.6 g
900	235.9 g	172.4 g	172.4 g

Drehzahl (min^{-1})	Head A ($\phi=52$ cm) g force	Head B ($\phi=38$ cm) g force	Head C ($\phi=38$ cm) g force
910	241.1 g	176.2 g	176.2 g
920	246.5 g	180.1 g	180.1 g
930	251.9 g	184.1 g	184.1 g
940	257.3 g	188.0 g	188.0 g
950	262.8 g	192.1 g	192.1 g
960	268.4 g	196.1 g	196.1 g
970	274.0 g	200.2 g	200.2 g
980	279.7 g	204.4 g	204.4 g
990	285.4 g	208.6 g	208.6 g
1000	291.2 g	212.8 g	212.8 g
1010	297.1 g	217.1 g	217.1 g
1020	303.0 g	221.4 g	221.4 g
1030	308.9 g	225.8 g	225.8 g
1040	315.0 g	230.2 g	230.2 g
1050	321.0 g	234.6 g	234.6 g
1060	327.2 g	239.1 g	239.1 g
1070	333.4 g	243.6 g	243.6 g
1080	339.7 g	248.2 g	248.2 g
1090	346.0 g	252.8 g	252.8 g
1100	352.4 g	257.5 g	257.5 g
1110	358.8 g	262.2 g	262.2 g
1120	365.3 g	266.9 g	266.9 g
1130	371.8 g	271.7 g	271.7 g
1140	378.4 g	276.6 g	276.6 g
1150	385.1 g	281.4 g	281.4 g
1160	391.8 g	286.3 g	286.3 g
1170	398.6 g	291.3 g	291.3 g
1180	405.5 g	296.3 g	296.3 g
1190	412.4 g	301.3 g	301.3 g
1200	419.3 g	306.4 g	306.4 g

Example:

A centrifuge with an effective radius of 260 mm necessitates an rpm of 1100 in order to reach the required RCF of 350 g.

Safety centrifuge for fat content determination

3680-L according to Roese-Gottlieb's method

SuperVario-N

3680 multi-purpose centrifuge for all butyrometers
see p. 48 for more details



Accessories for SuperVario-N

Head A

centrifuge head for a maximum
of 36 butyrometer buckets or 18 Babcock buckets
Radius of head A: 260 mm

3685

Accessories:

Butyrometer bucket: art. no. 3631, p. 46

Babcock bucket: art. no. 3632, p. 46



Head B

centrifuge head (protective tank)
for a maximum of 8 Mojonnier tubes
Radius of head B: 190 mm

3686

Accessory:

Mojonnier tubes: art. no. 3870, 3871, p. 55



Head C

centrifuge head
for a maximum of 6 buckets for solubility index tubes
Radius of head C: 190 mm

3687

Accessories:

Bucket for solubility index tubes: art. no. 3633, p. 46

Solubility index tube (ADPI glass): art. no. 3634, p. 46



WB-436 D universal water bath (digital)

digital temperature display (actual value)
digital set-point temperature control
PT 100 sensor (platinum sensor)
stop watch (1 to 99 min. with acoustic signal)

stainless steel inner and outer casing
external heating: heating elements are located separately in the casing
protection against overheating (even when tank is empty)
use with distilled water preferable

Technical specifications:

Temperature range: up to 100°C
Connection values: 230 V / 50 Hz ... 60 Hz / 1000 W
Dimensions (l x w x h): 396 mm x 331 mm x 265 mm
Volume: 16 l
Weight: 10 kg

3707 without butyrometer stand (art. no. 3717)



WB 436-A universal water bath (analogue)

Like art. no. 3707 but with analogue temperature control (adjusting knob), temperature display with thermometer (included in the scope of delivery), thermostatic heat controller

Stainless steel inner and outer casing
External heating: heating elements are located separately in the casing
Protection against overheating (even when tank is empty)
Use with distilled water preferable

Technical specifications:

Temperature range: up to 100°C
Connection values: 230 V / 50 Hz ... 60 Hz / 1000 W
Dimensions (l x w x h): 396 mm x 331 mm x 265 mm
Volume: aprox. 16 l
Weight: 10 kg

3708 without butyrometer stand (art. no. 3717)



Accessories for water baths WB 436 (art. no. 3707, 3708)

3717

Butyrometer stand for WB-436
stainless steel, for 36 butyrometers



3718

Mojonnier stand
stainless steel
for 10 Mojonnier tubes



3727

Universal shelf
stainless steel



3737

Reductase insert
for 99 samples



3747

Lid for Reductase test



3754

“Delvo-Test” insert



3766-G

Butyrometer tubes, closed
brass, for butyrometer stand (art. no. 3717)

3766-O

Butyrometer tubes, open (for art. no. 3707, 3708)
brass, for butyrometer stand (art. no. 3717)

Safety reading lamp

for safely and precisely
reading of butyrometers

anti-glare illumination, lens with protective Plexiglass cover,
adjustable height and lens distance, cord-operated switch

3800 230 V / 50 ... 60 Hz



Shaking machine

for extraction tubes according to Mojonnier's method

for forceful, uniform and reproducible mixing,
230 V / 50 ... 60 Hz

3850 for 4 Mojonnier tubes

3851 for 6 Mojonnier tubes

Shaking machine

for 36 butyrometers with stand

3852 230 V / 15 Watt, 915 x 270 x 300 mm (l x w x h)

Extraction tube with rounded bulb

acc. to Mojonnier's meth. with cork stopper (art. no. 3872)
suitable rubber stopper (art. no. 3310)

3870

Extraction tube flattened bulb

acc. to Mojonnier's meth., with cork stopper (art. no. 3872)
suitable rubber stopper (art. no. 3310)

3871

Cork stopper for extraction tube

according to Mojonnier's method (art. no. 3870, 3871)

3872

Wooden stand

for 12 extractions tubes according to Mojonnier's method

3875



KJELDAHL'S NITROGEN DETERMINATION METHOD

Anna Politis, graduate engineer of nutrition technology and of biotechnology, reports

For more than 120 years, Kjeldahl's nitrogen determination method has been an internationally accepted standard. In 1883, chemist Johan Kjeldahl developed this method for the quantitative determination of nitrogen. In the dairy industry, Kjeldahl's nitrogen determination method is used to determine protein content. The process is applied in accordance with ISO, DIN 8968-2/8968-3. To calculate the protein content, the nitrogen value obtained is multiplied by the product-specific factor for milk and dairy products, **6.38**. Nowadays, the entire determination can be automatised, even with multiple samples.

KJELDAHL'S NITROGEN DETERMINATION METHOD

1. Principle

The sample is solubilised with concentrated sulphuric acid and potassium sulphate in the presence of the catalyst copper sulphate. This causes the nitrogen bonded in the organic compounds to convert into the inorganic compound ammonium sulphate. Boiling the solution with sodium hydroxid releases ammonia from the ammonium sulphate. This is then led together with water vapour through a distillation apparatus. An ammonia-water solution results, which is introduced into a exactly defined amount of boric acid solution. By acidimetric titration, the amount of bonded boric acid and the nitrogen content is determined. Then protein content of the sample can be calculated with the help of the protein specific conversion factor.

2. Chemicals needed

- 2.1 Potassium sulphate (K_2SO_4) with a low nitrogen content.
- 2.2 Copper sulphate solution ($CuSO_4 \cdot 5H_2O$): 5.0 g copper sulphate pentahydrate are dissolved and stirred into 100 ml of water.
- 2.3 Sulphuric acid: with 98 % by weight, nitrogen free, $\rho_{20}(H_2SO_4) \sim 1,84$ g/ml
- 2.4 Sodium hydroxid: with a low nitrogen content and a percent by weight of 30 g of sodium hydroxide per 100 g.
- 2.5 Indicator solution: 0.1 g of methyl red is dissolved into 95 % ethanol and diluted to 50 ml with ethanol. 0.5 g of bromcresol green are dissolved into 95 % ethanol. One part methyl red solution is mixed with five parts bromcresol green.
- 2.6 Boric acid solution (H_3BO_3): 40.0 g of boric acid are dissolved in one litre of hot water. The solution is cooled and the volume is adjusted to one litre. 3 ml of indicator solution (2.5) are added and the solution is stirred and stored in a borosilicate glass bottle (the solution is light yellow). The solution must be kept away from light and ammonia vapour during storage.
- 2.7 Hydrochloric acid: the concentration must be 0.1 ± 0.0005 mol/L.
- 2.8 Ammonium sulphate [$(NH_4)_2SO_4$]. Before use, the ammonium sulphate must be dried for at least two hours at $102 \pm 2^\circ C$ and cooled in a dehydrator to room temperature. The purity of the dried substance must equal 99.9 %.
- 2.9 Water: distilled or fully desalinized water or water of equivalent purity.
- 2.10 Sucrose: with a nitrogen content of less than 0.002 %.
- 2.11 Tryptophan or lysine hydrochloride with a purity of at least 99 %.

3. Equipment and appliances

- 3.1 Laboratory scale: suited for weighing increments of 0.1 mg.
- 3.2 Boiling stones: grain size 10; (boiling stones not for reuse).
- 3.3 Water bath, adjustable to $(38 \pm 1)^\circ\text{C}$, suited for tempering milk and dairy products.
- 3.4 Digestion apparatus (art. no. 4200) consisting of a metal block, equipped with a heater and temperature regulator, and an emissions collector (supply voltage 230 V, temperature range up to 450°C).
- 3.5 Extraction apparatus with suction system (Behrosog 3, art. no. 4203): this neutralises dangerous vapours.
- 3.6 Distillation apparatus suitable for connecting to the 250 cm³ digestion flask (art. no. 4210).
- 3.7 Titrator: an automatic titrator or a burette with a nominal volume of 50 ml and a scale division calibre of at least 0.1 ml in accordance with ISO 385, class A (art. no. 4220).
- 3.8 Digestion flask with a nominal volume of 250 cm³
- 3.9 Pipette: suited for dispensing 1 ml of copper sulphate solution (2.2).
- 3.10 Erlenmeyer flask, nominal volume of 500 ml.


4. Preparation

The milk sample is heated to $38 \pm 1^\circ\text{C}$ in a water bath, stirred gently, and then cooled to room temperature. $(5 \pm 0,1)$ g are weighed out to exactly 0.1 g into a digestion flask.

5. Procedure

1 Digestion

Total solubilization time: 1.75-2.5 hours. **Note: must be executed under a flue.**



12 g of potassium sulphate, 1 ml of copper sulphate solution (2.2), about $(5 \pm 0,1)$ g of the heated and mixed milk sample and 20 ml of sulphuric acid are introduced into a digestion flask (the exact amount of milk must be determined to ± 0.1 mg and recorded because it will later be the basis for the nitrogen calculation. See "calculation"). The digestion flask is mixed carefully.

A suitable temperature program is selected on the digestion device, and then the digestion flask is set on the heating block. A glass emissions collector is carefully fixed onto every digestion flask. The entire apparatus is connected to a second apparatus (Behrosog 3, art. no. 4203) by a hose which can neutralise dangerous vapours. Selecting the following temperature program is recommended:

- 1) Preheat the heating block at 200°C for 10 min.
- 2) Heat the sample to 200°C for approx. 30 min.
- 3) Continue at 420°C for approx. 90 min for digestion (at a capacity of $10^\circ\text{C}/\text{min}$).

The digestion time is to be adjusted so that the maximum nitrogen contents are obtained. Too short or too long digestion times can lead to low values.

After digestion, the samples are removed from the heating block and cooled at room temperature for 25 minutes. Afterwards, they are put into the distillation apparatus.


- The addition of potassium sulphate serves to increase the boiling temperature of the sulphuric acid and the addition of copper sulphate serves as an oxidation catalyst. They are also available as Kjeldahl tabs (art. no. 4230/4231). If the process is executed with tabs, then 5 ± 0.1 g of milk are mixed with 20 ml of sulphuric acid and 2 Kjeldahl tabs and left to sit for 5 min. Then the temperature program can be carried out.
- During the heating of the sample, foam is not allowed to climb higher than 4 - 5 cm below the flask opening.
- The sulphuric acid is added in such a way that any copper sulphate solution, potassium sulphate or milk which may have adhered to the flask neck is flushed down. If the flask is sealed airtight, it can also be stored for later digestion.
- To determine the specific digestion time, it is advisable to execute preliminary tests with samples high in protein and fat.
- Substantial crystallisation is a sign of too little sulphuric acid and can lead to low protein values. It is therefore advisable to reduce the loss of sulphuric acid by minimising the amount sucked.
- Before the hot digestion flasks can be taken from the digestion block, it must be ensured that no condensed fluid has collected in the extraction apparatus. If it has, the suction volume must be increased and the condensed liquid removed.
- The undiluted digestion should not be stored in the flask for a long period of time (overnight) for any reason. There is a risk that the sample will solidify and it is difficult to bring it into solution form again. After the sample is cooled down and diluted with 70 ml of water, there is no problem to keep it overnight.

KJELDAHL'S NITROGEN DETERMINATION METHOD

Distillation

2

Total distillation time: 5-7 min



Do not touch any parts of the distillation apparatus either during a distillation or for a while after.

They could be hot.

The amount of water and sodium hydroxid, the reaction time, the distillation time, the heat output of the vapour generator and the suction time for the remains distilled out of sample are programmable. With the push of a button, the desired menu is found and by pushing the button again it is selected. Another push changes the value and a final push saves it.

The distillation apparatus must be degassed if it has not been used for a long time or if it is being used for the first time. To do this, "options", and then "direct input" are selected. Then, "H₂O in sample" is selected and the knob is held down until water runs into the digestion apparatus. Next, "NaOH" is chosen and the button is pushed down until soda lye runs into the digestion vessel. Finally, the menu "extractionPro" is selected by pushing "next" and the operating knob is pushed down until the chemicals are sucked from the digestion vessel. The aeration process is now finished.

A trial run without a sample must be executed daily before beginning the distillation process. To do this, a discharge hose is introduced into an empty digestion vessel, the main menu "options" and then "direct input" are selected. Then, "vapour" is chosen. By quickly pushing the operating button again, the vapour discharge is begun. Renewed pushing of the button ends the process. The process should not be ended until there is 1 cm of distillate in the Erlenmeyer flask. Finally, "extractionPro" is selected and the operating button is pushed down until the water is extracted from the digestion vessel.

After degasing and the test run, the distillation of the sample is carried out. A 500 ml Erlenmeyer flask for the boric acid solution is placed under the distillation apparatus outlet pipe. With the "start" menu option, the distillation process is begun. It is recommendable to run the following distillation program:

Bidest water: 70 ml (5 sec)
NaOH: 70 ml of a 30 % solution (7 sec)
Distillation time: 5 min
Vapour capacity: 90 %
Sample suction: 30 sec.
Boric acid addition: 50 ml (4 sec)

The vapour distillation is begun and the ammonia, released through the addition of sodium hydroxid, is distilled with vapour. The distillate is absorbed in the boric acid solution (2.6). After the distillation program terminates, the digestion flask is removed from the apparatus, the distillation hose is flushed with distilled water and the Erlenmeyer flask containing the sample is also removed from the apparatus.

■ The water distiller must be set up on a stable laboratory bench with an even, horizontal support which is located near a cold water supply and a drain. The water pressure must be at least 0.5 bar.


■ Before starting the operation, all hoses must be connected and the coolant engaged. The storage tank must be correctly positioned and the fluid level checked. The water vapour discharge hose must be introduced into the digestion flask. The water distiller is equipped with a safety gate.

■ During the first distillation, the water vapour comes into contact with cold pipes and glass parts. This leads to increased build-up of condensation which can in turn lead to excessive sample dilution and liquid volume in the digestion vessel. A trial run is therefore essential. The discharge of water vapour with a temperature of approx. 106°C creates loud noises. This is no cause for alarm.

■ The distillation is conducted until a distillation volume of 150 ml is obtained.

■ About 2 minutes before the end of the distillation, the Erlenmeyer flask is lowered in such a way that the end of the discharge pipe is no longer submerged in the acid solution. The pipe must be flushed with water. This water is collected in the Erlenmeyer flask.

3 Titration 1 min



The boric acid absorption solution (which contains the indicator) is titrated with 0.1 M standard hydrochloric acid. The hydrochloric acid is added until the first trace of pink coloration arises. The volume of the hydrochloric acid consumed is read off on the burette at 0.05 ml. An illuminated disc can serve as a neutral background and enables the user to determine the colour change accurately at the end of the titration.

- The mixture of methyl red and bromcresol green (see 2.5, 2.6) serves as the indicator. The indicator is responsible for the colour change and signals the end of the titration.
- The neutral background improves the accuracy and reproducibility of the results. This means that the titrations are always carried out under optical conditions that are as similar to each other as possible.

Blind trial

The blind trial is carried out in the same way as above. The sample is replaced with 5 ml of water and 0.85 g of sucrose. The volume of hydrochloric acid consumed during titration is recorded.

- The blind trial is important for the calculation of the nitrogen content of the sample.

Calculation and analysis

The nitrogen content, given in g of nitrogen per 100 g of the product, is calculated using the following numerical equation:

$$W_n = \frac{1,4007 (V - V_0) C_s}{W_t}$$

W_n: the nitrogen content of the sample

V: the volume of hydrochloric acid consumed during titration of the sample

V₀: the volume of the hydrochloric acid consumed during titration of the blind trial (see blind trial)

C_s: the exact molarity of the hydrochloric acid, given to four decimal places

W_t: the mass of the test sample in grams, given within 0.1 gram

To calculate the protein content of the sample, the W_n value must be multiplied by 6.38.

Model calculation

If the Kjeldahl determination yields a nitrogen content of 55 %, a protein content of 3.5% results (55% x 6.38).

- Kjeldahl's digestion method is not specific to amino acids and proteins and includes all organically bonded nitrogen. Other non-protein compounds are also digested and collected (NPN: non-protein nitrogen). However, the proportion of these compounds is very small and is disregarded in the calculation.

- If the non-protein containing nitrogen should also be established, then method must be executed in accordance with DIN EN ISO 8968-4. If only the protein nitrogen should be determined, then the milk proteins must first be separated. 5±0.1 ml of milk diluted with 5±0.1 ml of water is washed in stages with in total 60 ml of 15 % (w/v) trichloroacetic acid in accordance with DIN EN ISO 8968-5, the proteins are precipitated and finally filtered out into a hard paper filter. The filtrate contains the components of the non-protein nitrogen and the filtered-out precipitate contains the protein nitrogen. The filter with the precipitate is put into a digestion vessel and Kjeldahl's nitrogen determination method is carried out as described above. The protein content is calculated by multiplying by a factor of 6.38.

- The value 6.38 is specific to milk and dairy products and was established because milk proteins have a nitrogen content of 15.65 % (100:15.65 = 6.38).

KJELDAHL'S NITROGEN DETERMINATION METHOD

Alternatives, rapid methods

There is a faster way to execute Kjeldahl's method than the standard process. In accordance with ISO 8968-3, smaller milk samples are used (2 g, exactly weighed within 1 mg). 2 g of the sample, a catalyst tablet (consisting of 5 g K_2SO_4 , 0.105 g $CuSO_4 \cdot 5H_2O$ and 0.105 g TiO_2), 10 ml of 98% sulphuric acid, and a few drops of antifoam agent (30% silicon compound) are introduced into a 250 ml digestion flask. The contents are mixed carefully, 5 minutes are allowed to pass and then 5 ml of 30 % hydrogen peroxide are added at the side of the flask. The sample is left to sit for 10-15 minutes before digestion. The heating block is preheated to 400°C for 10 minutes and the sample is heated at 400°C for 60 minutes. Afterwards, the sample is cooled to room temperature and diluted with 50 ml of water. 55 ml of 30 % NaOH and 50 ml of 4 % boric acid are used to absorb the distillate. HCL 0.05 M is used for titration. 0.2 ml of a mixture of 0.03 % methyl red and 0.17 % bromcresol green in 95 % ethanol serves as the indicator.

■ *Caution! The addition of hydrogen peroxide causes an intense reaction.*

■ *The blind trial is executed with 2 ml of water and 0.25 g of sucrose. The effectiveness of the decomposition is tested with 0.08 g of tryptophan or 0.06 g of lysine hydrochloride.*

Monitoring the process

The accuracy of the process should be reviewed regularly. A loss of nitrogen should be tested for. As a sample, 0.12 g of ammonium sulphate $[(NH_4)_2SO_4]$ and 0.85 g of sucrose are used here. Kjeldahl's method is carried out under the same conditions as with normal samples and the percentage of nitrogen content must be between 99.0 % and 100 %. The determination with ammonium sulphate is of use in detecting nitrogen losses during decomposition or distillation and differences in concentrations in the titration agent.

■ *In order to monitor the effectiveness of the decomposition, 0.18 g of tryptophan or 0.16 g or lysine hydrochloride and 0.67 g of sucrose are used. 98 % of the nitrogen content must be recovered. If that is not the case, either the decomposition temperature or time is insufficient or the sample is charred.*

Determining the nitrogen content of dairy products

The reference standard process can also be adapted for other milk products. The only difference in the method is the sample amount. The sample must show a protein amount of 0.15-0.30 %. It is therefore advisable to use the following sample amounts:

Condensed milk:	3 g
Powdered skim milk:	1 g
Whey:	10 g
Cream cheese:	3 g

The method is carried out in the same way as the standard process.

Kjeldahl digestion apparatus K8

Heating block and glass extraction system for 8 samples, for connecting to the Behrosog suction station. Suitable for 250 ml digestion flasks. The front side of the sample rack is covered. Stable and robust construction. The block casing as well as the extraction hood is made of acid-proof, rust-free stainless steel.

Programmable for up to 10 different temperature steps.
Maximum temperature 450°C,
time adjustment range 0-999 min
230V, 50 Hz, weight: 28 kg

4200 480 x 510 x 765 mm (w x d x h)



Digestion flask

4201 250 ml



Suction station Behrosog 3 with cooler

Extracts aggressive acid vapours during digestion. In the process, an upstream two-staged pre-filter edulcorates and precipitates the toxic substances.

230 V, 50 /60 Hz,
weight: 18 kg
suction pump: 40 l/h

4203 80 x 340 x 400 mm (w x d x h)



Kjeldahl distillation apparatus S-3

behind a safety screen,
automatic water vapour production,
manual or automatic addition of H_2O , NaOH .
programmable distillation time and reaction time,
automatic suction of sample remains,
automatic fluid level surveillance of the storage tank.

230 V, cooling water usage: 3 L/min,
weight: 35 kg

4210 410 x 675 x 410 mm (w x d x h)



Automatic titrator STI

The titration station consists of a burette with a digital display and a magnetic stirrer with a custom-fit holder for an Erlenmeyer flask. Result accuracy and reproducibility are enhanced due to a viewing shield which serves as a neutral background.

230 V, 50 / 60 Hz, weight: 3.5 kg

4220 330 x 200 x 600 mm (w x d x h)



Kjeldahl Tabs KT1

consisting of 5 g of potassium sulphate,
0.5 g of copper sulphate

4230

Kjeldahl Tabs KT2

consisting of 5 g of potassium sulphate, 0.105 g of copper sulphate, 0.105 g of titanium dioxide

4231

pH VALUE MEASUREMENT

Anna Politis, graduate engineer of nutrition technology and of biotechnology, reports

The pH value is a measurement of the H^+ activity. Simply put, it is a measurement of the concentration of acid ($pH < 7$) or base ($pH > 7$). The formal definition was formulated by chemist Soerenson:

$$pH = -\log a_{H^+}$$

The pH value is the negative decadal logarithm of the activity of protons a_{H^+} in mol/l. This value is measured by means of a pH meter with a suitable electrode in accordance with DIN 38404-C5. pH meters measure the potential difference between measuring electrodes and reference electrodes. According to the Nernst equation, the potential difference changes by 59 mV per pH unit. The pH meter must be calibrated regularly from time to time. The calibration is carried out by means of standard buffer solutions with defined pH values.

During a 2 point calibration, the zero point is set by an additive correction with the buffer solution ($pH=7$). Subsequently, the end value (e.g. $pH=4.01$) is set over a multiplicative correction (transconductance). How often the meter should be calibrated depends on how exact the measurements should be and varies from meter to meter. If the pH meter is not in constant use, it should be calibrated before every measurement. If the electrode is continuously in the storage liquid, it doesn't have to be calibrated as frequently. For each calibration, 20 ml of buffer solution are needed. Solutions should not be used more than once. The bottles containing the buffer solution must be sealed immediately after use. Alkaline calibration solutions are more sensitive. Their pH values change because they absorb CO_2 from the air. A sealed bottle of buffer solution keeps for several months to two years. Between calibrations or measurements, the sensor must be rinsed with distilled water but not wiped off. Excess drops can be dabbed away with a soft cloth.

The pH value is temperature-dependent; therefore the temperature must always be recorded along with the entry of the pH value. Nowadays most pH meters are equipped with a temperature measuring unit. This allows the temperature influence to be compensated during the measurement.

A pH meter requires regular maintenance to measure optimally.

With electrodes with refillable electrolyte, the fluid level of the electrolyte solution must be checked. The level of the reference electrolyte must always be a few cm above the fluid level of the measurement solution. If necessary, the KCl solution 3M must be refilled after removing the seal over the filler hole. During use, the refill hole for KCl should always be open, otherwise the solution cannot be diffused out. If the electrode is no longer needed, it should be quickly rinsed, the refill hole for KCl should be sealed and the electrode should be stored in the 3M KCl solution so that it doesn't dry out.

During transport and storage, KCl solution can leak out of the protective cap, out of which crystallised white potassium chloride is formed. This salt deposit has no effect on measuring accuracy and can be easily washed off with water. If the electrode is dried out, then it must be soaked in 1ML HCl and subsequently reactivated for several hours in 3M KCl.

Should constant deviations be determined during measurements, the electrode must be checked for possible contamination. Depending on the type of contamination, different cleaning measures are recommended.

- *To clean off fat or oil deposits, the membrane must be degreased with cotton which has been soaked in acetone or soap solution.*
- *If protein has settled on the diaphragm, the electrode is soaked in HCl/Pepsin solution for approx. 1-2 hours.*
- *In case of a silver sulphide contamination, the electrode is to be set in a thiocarbamide solution and left to soak.*
- *To remove inorganic films, the electrode is dipped into 0.1 M HCl or 0.1 M NaOH. With 40-50°C solutions, cleaning is more effective.*
- *After every cleaning procedure, the electrode is to be set in a 3M KCl solution about a quarter of an hour a new conditioning and subsequently calibrated once again.*



Battery/pocket pH meter



Laboratory pH meter

pH meters

Laboratory pH meter

electrode not included, with DIN electrode connection, for compatible single rod electrodes, see art. no. 4336

Knick 766 easy to use measurement device for pH, mV and °C: electrode adjustment and monitoring self-test, automatic temperature compensation recorder output, calibration data memory

4310



Knick 765 plus RS 232 interface for computer and printer (GLP documentation)

4311



Battery/pocket pH meter

electrode not included in the scope of delivery (see art. no. 4370, 4380)

Knick 911 highly developed measuring device for pH, mV and °C with support for use on a table, protected from dust, water, as well as impact: automatic calibration, buffer recognition, temperature compensation, self-test, DIN electrode connection

4315



Knick 913 similar to 911, but with additional measurement value storage and interface for computer and printer (GLP documentation), with DIN electrode connection

4317



Pt 1000 temperature sensor

for Knick 911, 913 (art. no. 4315, 4317),
with DIN plug

4319**Single rod electrode SE 100**

with integrated temperature sensor Pt 1000
compatible with Knick 766, 765 (art. no. 4310, 4311),
with DIN plug

4336**Single rod electrode Inlab Basics**

suitable for milk and other fluids
fixed cable with DIN plug

4350**Insertion electrode Inlab Solids**

insertion head electrode
with cable and DIN plug

4360**Insertion electrode Inlab Solids**

without cable

4361**Single rod electrode SE 104**

for insertion measurements in cheese, meat, and sausage,
compatible with Knick 911, 913 (art. no. 4315, 4317)
fixed cable with DIN plug

4370**Single rod electrode SE 102**

integrated with temperature sensor Pt 1000
design compatible with Knick 911/913 (art. no. 4315, 4317)
fixed cable with DIN plug

4380**Buffer solutions**

in 250 ml PE bottles

4390

pH 4.01

4391

pH 7.00

4392

pH 9.21

KCl solution

in 250 ml PE bottles
3 mol/L

4400

Cleaning agent for single rod electrode

in 250 ml PE bottles

4420 thiocarbamide solution for Ag-Cl diaphragms

Pepsin hydrochloric acid solution

4421 protein solvent

Reactivation solution

hydrofluoric acid

4422 in 25 ml PE bottles

pH Meter "pH 49"

in accordance with guideline 89/336/EWG

Battery type: 9 V

Operation temperature: 0-50°C

Electrode connection: pH / mV: BNC connector

4450 °C: DIN connector



Temperature sensor Pt 100

4451 for pH meter "pH 49"

pH single rod electrode EGA 184

4452 for pH meter "pH 49"

4453 pH single rod electrode with integrated Pt 100

4455 Platinum redox single rod electrode

4460 Buffer solution pH 4.01 / 250 ml

4461 Buffer solution pH 7.0 / 250 ml

4462 Buffer solution pH 9.18 / 250 ml



TITRATION APPARATUS

determination of acid content
to ascertain the degree of freshness

Titration apparatus STANDARD

complete with storage bottle, rubber stopper, burette with automatic zero point adjustment, sodalime with ascending tube, rubber pressure bulb, burette tip with pinchcock, one 1 ml and one 25 ml pipette, one 200 ml Erlenmeyer flask

4500 for milk: 0 - 25° SH

4501 for cream: 0 - 40 °SH

4510 for curd: 0 - 250 °SH
with porcelain mortar and pestle, 2 ml pipette
(without 1 ml and 25 ml pipette and Erlenmeyer flask)



Titration apparatus SIMPLEX

for milk and cream, complete with polyethylene bottle on a plastic base, burette with automatic zero point adjustment, precision titration by button control, one 1 ml and one 25 ml pipette, one 200 ml Erlenmeyer flask

4520 for milk: 0 - 25° SH

4521 for cream: 0 - 40° SH

Titration apparatus SIMPLEX

for general titration purposes, as above,
but without accessories

4530 with burette 0 - 10 ml: 0.05

4540 with burette 0 - 25 ml: 0.1

4550 with burette 0 - 50 ml: 0.1



Titration apparatus

with bottle and holder
without accessories

4654 0 - 100° Dornic

4655 0 - 40° Dornic

Protein titration apparatus

with storage bottle,
for use with 25 ml of milk,
special burette with automatic
zero point adjustment,
sodalime with ascending tube, rubber bulb,
outlet tip, pinchcock,
one 1 ml, one 5 ml and one 25 ml volumetric pipette,
two 250 ml short beakers,
two 1 ml: 0.01 measuring pipettes

4660 0 - 6 ET: 0.02

Acidity tester

for determining the degree of freshness
of unpasteurised milk

4705



Salt test

for butter and cheese
see art. no. 4530, 4540,
but with brown storage bottle

4760 for butter 10 ml: 0.05

4770 for cheese 25 ml: 0.1

SEDILAB sediment tester

easy-to-use manual sediment tester,
with clamp for attaching to tables, stainless steel

4800 for 500 ml of milk




SEDILAB-E sediment tester

for serial testing of liquids for particle contamination,
particularly for sediment testing of milk,

splash-proof design, approx. 800 samples per hour,
sharply defined sediment images,
220 V / 50 Hz

4810 for 500 ml of milk



ASPILAC sediment tester

pump design for direct suction from a can,
Plexiglass casing for original filter papers

4905 for 500 ml of milk


Filter papers

4910 with area for records, 1000 pieces, Ø 28 mm, 80 x 45 mm



Filters, round

4911 32 mm, 1000 pieces



Reference table

4920 with 3 purity grades, German standard



Pipetting syringes

for determining out nutrient and dye solutions, self-priming, can be sterilised

5110 adjustable to 1 ml

5111 adjustable to 2 ml

5112 adjustable to 5 ml



Methylene blue tablets

for bacteria count estimation
50 pieces

5140

Resazurine tablets

for LOVIBOND comparator (art. no. 5160), 100 pieces

5150

LOVIBOND comparator 2000

for resazurine tests,
housing for 2 test tubes
for colour comparisons,
with milk observation stand,
without colour disc (see art. no. 5161)

5160**Colour disc**

for resazurine 4/9 with 7 standard reference colours

5161**Test tube**

set of 4 tubes

5162**Dry matter calculator**

according to Ackermann's method, for milk


5360**Butter melting beaker**

for the determination of water content in butter

5400 Aluminium, 30 g; l = 51 mm, d = 60 mm

5401 Aluminium, 50 g; l = 66 mm, d = 64 mm

5420 Tongs

**Glass stirrer**

pestle type, 140/6 mm

5430**Double-ended spatula**

pure nickel, 150 mm

5440**Butter testing spoon**

Plexiglass

5450

Crystalline quartz sand

0.6-1.2 mm grain size, calcined quality

5460 washed, 1 kg, transport costs available on request

5461 washed, 3 kg, transport costs available on request

5462 washed, 5 kg transport costs available on request

5463 washed, 25 kg transport costs available on request

5464 washed, 10 kg transport costs available on request

Aluminium foil

5470 150 x 190 mm, 1000 pieces

Weighing dish

aluminium, with lid, (numbered on request)

5490 75 x 30 mm



Bunsen burner

for propane gas
(other gas types available on request)

5550

Infrared burner, up to 750°C
suitable for fast, contact-free heating

5571 0.9 kg, 100 x 100 x 100 mm

5572 Output regulator



Water paper

Indicator paper for determining moisture distribution in butter

5600 40 x 78 mm, 50 pieces

Test tube according to Beckel's method

for determining the acid value in butter

5601 5 ml / 11 ml, PE stand

Butter cutter

5605 wire gauge 0.5 mm

Separating funnel

for extraction

5606 250 ml

**Thin layer chromatography chamber**

5607 200 x 200 mm

**Thin layer chromatography plates**

25 silicon dioxide gel plates
with aluminum liner
can be cut with scissors

5608 200 x 200 mm

Pocket refractometer

for measuring the degree of evaporation in milk and determining the concentration in various fields of application, inc. case the internationally approved Brix scale enables the weight percentage of dry matter to be determined directly.

5610 0 – 32 % Brix: 0.2 % for milk, fruit juices, soft drinks

5612 28 – 62 % Brix: 0.2 % for concentrated fruit juice

5613 45 – 82 % Brix: 0.5 % for honey



Digital hand refractometer

can be switched from 1.330 – 1.5318 n_D :
resolution 0.1 % Brix, 0.0001 n_D
automatic temperature compensation from 10 - 40°C

5614 0 - 95 % : 0.1 % Brix



Digital Abbe refractometer

LED display 590 nm, serial interfaces RS-232
and RS 422, 115/230 V, 50/60 Hz
1.3000 – 1.7200 n_D : 0.0001 n_D
5 kg - 140 x 275 x 300 mm

5620 0 - 95 %: 0.1 % Brix, 0 - 99°C: 0.1°C

Humidity measuring device MLB 50-3

for the fully automatic determination of moisture content or dry substance content
data interface RS 232

5670 5.5 kg - 217 x 283 x 165 mm



Accessories for humidity measuring device MLB 50-3

Aluminium specimen dish

5671 92 mm diameter, packs of 80 pieces

Circular glass fibre filter

5672 for splashing or caking specimens

Matrix needle printer

5673

Specimen dish

aluminium

5674 100 x 7 mm, pack: 100 pieces

Reference drier RD-8

for determining the moisture content of powdered milk in accordance with ISO/DIN 5537, IDF 26

8 samples can be dried simultaneously under precisely defined conditions (87°C / 33 ml/Min airflow).

Connections: a) 230 V / 115 V, 520 W
b) 2.5 bar ... 7.5 bar
Temperature range: adjustable, up to 110.0°C
Stability: +/- 0.3°C

5700



Accessories for reference drier RD-8

specimen container

PP, 20 pieces


5701



Lid for specimen container

PP plastic, 20 pieces

5702



Cap

PP plastic, 20 pieces


5703



Filter

100 pieces

5704



Loading arm

for easy and exact positioning of the filter
in the specimen receptacle, acrylic

5705**Weighing stand****5706****Stand for lids and caps****5707****Flowmeter**

for measuring the air flow
in the reference drier RD-8
ADM 1000

5708**Round aluminium foil**

130 x 0.03 mm, 1000 pieces

5712

Analytical scale

with modern all-glass wind protection,
automatic internal adjustment every 3 hours
or with a temperature change of $>0.8^{\circ}\text{C}$
display change from piece to weight,
GLP/ISO logging possible,
percentage determination, RS 232 data interface,
under-floor weighing possible,
gauge or calibration certificate for additional charge

weighing plate diameter: 85 mm

5810 160 g: 0.1 mg

5811 220 g: 0.1 mg



Precision scale

with formulation memory, unit counter,
GLP/ISO logging possible,
percentage determination, RS 232 data interface,
under-floor weighing possible

weighing plate: 130 x 130 mm

5820 1600 g: 0.01g



Further scales available on request

HEATING CABINETS UNB

with natural air circulation for standard tempering tasks from 30-220°C

Order no.	Model	Volume (litres)	ext. dimensions w/h/d (mm)	int. dimensions w/h/d (mm)	supporting ribs/ slide-in plates	Watts/ volts	Kg net	Equipment type/fixtures
6000	UNB 100	14	470/520/325	320/240/175	2/1	600/230	20	Digital (switch-off) clock 99 hours 59 min.
6001	UNB 200	32	550/600/400	400/320/250	3/1	1100/230	28	
6002	UNB 300	39	630/600/400	480/320/250	3/1	1200/230	30	

HEATING CABINETS UFB

with forced air circulation for standard tempering tasks from 30-220°C

6008	UFB 400	53	550/680/480	400/400/330	4/2	1400/230	35	Digital (switch off) clock 99 hours 59 min.
6009	UFB 500	108	710/760/550	560/480/400	5/2	2000/230	50	

INCUBATORS INE

with natural air circulation for tempering tasks from 30-70°C

Order no.	Model	Volume (litres)	ext. dimensions w/h/d (mm)	int. dimensions w/h/d (mm)	supporting ribs/ slide-in plates	Watts/ volts	Kg net	Equipment type/fixtures
6035	INE 200	32	550/600/400	400/320/250	3/1	1100/230	28	Excellent Fuzzy PID controller with two integrated clocks (running time 1 min. to 999 hours and weekly program timer) and triple thermal safety fuse, air turbine speed controller
6036	INE 300	39	630/600/400	480/320/250	3/1	1200/230	30	
6037	INE 400	53	550/680/480	400/400/330	4/2	1400/230	35	
6038	INE 500	108	710/760/550	560/480/400	5/2	2000/230	50	

STERILISING OVENS SNB

with natural air circulation for tempering tasks from 30-220°C

Order no.	Model	Volume (litres)	ext. dimensions w/h/d (mm)	int. dimensions w/h/d (mm)	supporting ribs/ slide-in plates	Watts/ volts	Kg net	Equipment type/fixtures
6047	SNB 100	14	470/520/325	320/240/175	2/1	600/230	20	Digital (switch-off) clock 99 hours 59 min.
6048	SNB 200	32	550/600/400	400/320/250	3/1	1100/230	28	
6049	SNB 300	39	630/600/400	480/320/250	3/1	1200/230	30	

REFRIDGERATED INCUBATORS WITH COMPRESSION COOLING ICP

for tempering tasks from 0-60°C

Order no.	Model	Volume (litres)	ext. dimensions w/h/d (mm)	int. dimensions w/h/d (mm)	supporting ribs/ slide-in plates	Watts/ volts	Kg net	Equipment type/fixtures
6070	ICP 400	53	558/967/486	400/400/330	4/2	500/230	68	PID process controller, serial and parallel interfaces, motorised inner air circulation
6071	ICP 500	108	718/1047/556	560/480/400	5/2	500/230	87	
6072	ICP 600	256	958/1335/656	800/640/500	7/2	700/230	144	

further instruments available on request

Laboratory furnaces

heating and incineration at up to 1100°C,
rust-free stainless steel furnace casing,
high-grade isolation, short heating-up time
230 V/50 Hz, 1.2 kW, Volume: 3L

6220 Internal dimensions: 160 x 140 x 100 mm,
External dimensions: 380 x 370 x 420 mm, 20 kg

Discharge viscometer

Easy-to-use viscometer for in-house measure-
ment of the viscosity of yogurt, curdled milk, kefir
and other products.

The stop time of the discharge of the measured
material serves as a measure of the viscosity.

6520 with stand and two different discharge nozzles

6521 Glass plate

6522 Stop watch

Visco tester VT6R Haake

rotation viscometer for measurements in accordance
with ISO 2555 and ASTM (the Brookfield method)

- measuring range 20 ... 13,000,000 mPas (cP)
- acoustic warning for measuring range
- RS 232C interface
- set of 6 spindles

6530 stand and case included in the scope of delivery



INHIBITOR DETECTION

6570 Delvotest SP-NT for 100 samples

6571 Delvotest plate test SP-NT
each for 96 samples

LACTODENSIMETER

Lactodensimeters are frequently used with an official calibration, or are officially calibrated with a certificate. Please refer to our price list or contact us for more information.

Lactodensimeter

for milk according to GERBER's method, large model,
negative scale, with thermometer in stem,
1.020 – 1.040: 0.0005 g/ml,
T = 20°C, 10 – 40°C, ca. 300 x 28 mm

6600

standard model

6602-E

officially calibrated, the calibrated range of the thermometer goes from 10°C to 30°C

6603-ES

officially calibrated, with certificate, the calibrated range of the thermometer goes from 10°C to 30°C



Lactodensimeter

for milk acc. to GERBER's meth., small model,
with thermometer in body,
1.020 – 1.035: 0.0005 g/ml,
T = 20°C, 0 – 40°C, ca. 210 x 17 mm

6610

standard model

6612-E

officially calibrated, the calibrated range of the thermometer goes from 10°C to 30°C

6613-ES

officially calibrated, with certificate, the calibrated range of the thermometer goes from 10°C to 30°C



Hydrometer

for milk in accordance with DIN 10290
without thermometer,
1.020 - 1.045: 0.0005 g/ml,
T = 20°C, ca. 350 x 25 mm

6620

standard model

6621-E

officially calibrated

6622-ES

officially calibrated, with certificate



Lactodensimeter

for milk according to Quevenne's method,
with coloured triple scale



6630 1.015 – 1.040: 0.001 g/ml, T = 20°C
with thermometer 0 – 40°C, approx. 290 x 22 mm

6630-15 1.015 – 1.040: 0.001 g/ml, T = 15°C
with thermometer 0 – 40°C, approx. 290 x 22 mm

6631 1.015 – 1.040: 0.001 g/ml, T = 20°C
without thermometer, approx. 210 x 22 mm

6631-15 1.015 – 1.040: 0.001 g/ml, T = 15°C
without thermometer, approx. 210 x 22 mm

Hydrometer for buttermilk serum

DIN 10293, without thermometer, T = 20°C,
1.014 – 1.030: 0.0002 g/ml, approx. 240 x 21 mm

6640 standard model

6641-E officially calibrated

6641-ES officially calibrated, with certificate

Buttermilk tester

according to Dr. Roeder's method
with thermometer in stem,
approx. 210 x 25 mm

6650 1.010 – 1.030: 0.001 g/ml, T = 20°C

Hydrometer for condensed milk

without thermometer, reading at top

6660 1.000 – 1.240: 0.002 g/ml, T = 20°C, approx. 310 x 19 mm

6661 1.040 – 1.080: 0.001 g/ml, T = 20°C, approx. 230 x 21 mm

Hydrometer for yogurt and chocolate milk

with thermometer incorporated in body,
reading at top
approx. 220 x 16 mm

6670 1.030 – 1.060: 0.001 g/ml, T = 20°C

Hydrometer for brine / Baumé

0 - 30 / 0.5 Bé, T = 15°C

approx. 240 x 17 mm

6680	without thermometer
-------------	---------------------

6681	with thermometer, 0 - 40°C
-------------	----------------------------

Hydrometer for boiler water

DIN 12791, M 100

without thermometer,

approx. 250 x 20 mm

6690	1.000 - 1.100: 0.002 g/ml, T = 20°C
-------------	-------------------------------------

Alcoholometer

0 - 100 Vol. %: 1.0, T = 20°C,

approx. 290 x 16 mm

6710	with thermometer
-------------	------------------

6711	without thermometer
-------------	---------------------

Hydrometer for amyl alcohol

DIN 12791, M 50

without thermometer

260 x 24 mm

6720	0.800 - 0.85: 0.001 g/ml, T = 20°C
-------------	------------------------------------

Hydrometer for sulphuric acid

DIN 12791, M 50

without thermometer

270 x 24 mm

6730	1.800 - 1.850: 0.001 g/ml, T = 20°C
-------------	-------------------------------------

6731	1.500 - 1.550: 0.001 g/ml, T = 20°C
-------------	-------------------------------------

Hydrometer

DIN 12791, M 50

for various liquids,

without thermometer, T = 20°C,

270 x 24 mm

6740	1.000 - 1.050: 0.001 g/ml
-------------	---------------------------

6741	1.050 - 1.100: 0.001 g/ml
-------------	---------------------------

6742	1.100 - 1.150: 0.001 g/ml
-------------	---------------------------

6743	1.150 - 1.200: 0.001 g/ml
-------------	---------------------------

Level gauge for lactodensimeters

internal diameter: 39 mm
length: 265 mm

6800

Stand

tripod with cardanic suspension
and hanging cylinder
for lactodensimeters art. no. 6610 - 6613

6810

265 x 60 mm

Replacement hanging cylinder

for art. no. 6810

6820

210 x 22 mm



Stand

with cardanic suspension,
overflow hanging cylinder,
compatible with all lactodensimeters
and hydrometers, incl. drip tray,
hoses and pinchcock

6830



THERMOMETER/ACCESSORIES

Dairy thermometer

with loop
special red filling
0 - 100°C: 1°C

7001 approx. 250 x 17 mm



Dairy thermometer

in plastic case with loop,
boil- and impact-proof, floatable,
special red filling
0 - 100°C: 1°C

7031 approx. 280 x 28 mm



Dairy thermometer

special red filling
replacement for art. no. 7031,

7041 approx. 250 x 17 mm

Universal thermometer

7046

special red filling
-10 to 100°C: 1.0, approx. 260 x 8 mm

Cooling chamber thermometer

7060

special blue filling
in plastic holder with loop and hook
-40 to +40°C: 1.0, approx. 200 x 20 mm

Control thermometer

7070-ES

special red filling
-10 to +100°C: 1.0, approx. 305 x 9 mm

7071

officially calibrated, with certificate
uncalibrated

Low temperature laboratory thermometer

7081

special red filling
-38 to +50°C: 1.0, approx. 280 x 8 x 9 mm

Maximum-minimum rod thermometer

7095

white coating, creosote filling
-35 bis + 50°C: 1.0, approx. 220 x 10 mm

7096

-10 bis + 100°C: 1.0, approx. 220 x 10 mm

The **psychrometer**

Measurement of relative humidity

A hair hygrometer is typically used to measure relative humidity. A strand of hair elongates when it absorbs moisture. The psychrometer functions more accurately. The instrument consists of two exactly matching thermometers (with as little deviation as possible). The mercury receptacle (alcohol is not used due to too high inaccuracy) of one thermometer is wrapped in a damp piece of absorbent cotton or the like. The other thermometer is kept dry and gives the temperature of the surrounding air. At a relative humidity of 100 %, both temperatures show the same temperature. If the humidity is lower than 100 %, the water on the "damp" thermometer evaporates. A lower temperature is shown on the damp thermometer as on the dry thermometer due to evaporation chill (the warmth necessary to evaporate is detracted from the thermometer and the piece of cotton). The humidity can be calculated from the temperature difference.

PSYCHROMETER

water storage tank,
2 calibratable thermometers with translucent glass
scale, with humidity table, lacquered wood plate

7100 -10 + 60: 0.5°C, approx. 190 x 12-13 mm

Polymeter (hair hygrometer)

for measuring relative humidity and temperature,
with scale for water vapour saturation pressure,
thermometer with special filling
Thermometer dimensions: approx. 130 x 12 mm
Hygrometer dimensions: Ø 100 mm

7110 Measuring range: 0 - 100 %, 0 - 30°C,

Humidity/temperature measuring device

with moisture sensor and NTC temperature sensor

Measuring range: -10 - +50°C, 0 - 100 % rH
Accuracy: $\pm 0.5^\circ\text{C}$, $\pm 2.5\%$ rH

7115



Digital thermometer 826-T4

contact-free measurement and core temperature
measurement in foodstuffs with one device

Measuring ranges:

contact-free / IR: -50°C to +300°C, accuracy: $\pm 2^\circ\text{C}$
with NTC sensor: -50°C to +230°C, accuracy: $\pm 1^\circ\text{C}$

7119

Digital second thermometer 926

for daily temperature measurements in the food
industry, for laboratory use
ISO calibration certificate for an additional price

Measuring range: -50 to +400°C: 0.1°C (1°C from 200°C),
Accuracy: $\pm 0.3^\circ\text{C}$.

7120



INSERTION/IMMERSION SENSORS

Robust **precision sensor**

7122 Ø 4 mm, length 110 mm

Stainless steel sensor for food

stainless steel ,

7123 Ø 4 mm, length 125 mm

Needle sensor

for quick measurements
without visible pinhole,

7124 Ø 1.4 mm, length 150 mm

Frozen goods sensor

screws in without pre-drilling

7125 Ø 8 mm, length 110 mm

TopSafe

protective cover against contamination,
water and impact

7127

FREEZING POINT DETERMINATION

One of the main focuses of Funke - Dr. N. Gerber Labortechnik GmbH
K. Schaefer, graduate engineer, W. Spindler, graduate physician report

HISTORY

The German chemist Beckmann, known for the thermometer named after him, began using the freezing point of milk in as early as 1895 to detect if it had been adulterated with water. The American Hortvet worked intensively with this method in 1920 and improved some of its essential features. The first thermistor cryoscopes were brought to the market in the 1960s. However, they had to be operated entirely by hand. At the beginning of the 1970s, the first automatic thermistor cryoscopes became available. With this development it was possible to determine the freezing point automatically at the push of a button.



A decisive step in the improvement of thermistor cryoscopy was displayed at the "FoodTec" tradeshow in 1984: Funke-Gerber introduced the first device with automatic calibration. This successful and intensive development work reached a new peak at the "Food-Tec" in 1988, where Funke-Gerber presented a fully automatic freezing point determination mechanism with a capacity of 220 samples per hour.

With the introduction of an indirect freezing point measurement device (e.g. LactoStar) for routine analysis, interest was focused primarily on reference devices which are able to determine freezing points in accordance with the applicable standards and regulations. These devices must satisfy the strictest requirements with regard to measuring accuracy. For this reason Funke-Gerber developed a programmable cryoscope with a resolution of 0.1 m °C. This instrument has proven its accuracy and reliability in countless laboratories all over the world. The product range has been expanded with a multi-sample device (CryoStar_{automatic}). Since January 2007, these instruments have been equipped with a graphic colour display. This makes it possible to show the entire freezing curve, in particular the process of the plateau search, with a patented screen presentation.

THE FREEZING POINT:

The freezing point of pure water is the temperature at which ice and water are in equilibrium.

If soluble components are added to this liquid, the freezing temperature lowers (becomes colder) because the ability of the water molecules to escape from the surface diminishes. As fat is not water soluble, it has no influence on the freezing point.

MEASURING PRINCIPLE:

The milk is cooled to -3°C (super-cooled) and crystallisation is induced by mechanical vibration. As a result of this freezing process, the temperature increases due to the released lattice energy and stabilises at a certain plateau which corresponds to the freezing point.

MEASURING PROCEDURE:

The freezing point of liquids is not just any temperature, but the exact temperature at which one part of the sample is in a solid state and another part is in a frozen state, whereby the parts are in equilibrium.

To measure the freezing point, the sample must therefore be brought into this state. In order to do this, a certain procedure must be followed, which is carried out in the following way:

First the sample must be cooled to under the actual freezing point while being stirred. Stirring is necessary for 3 reasons:

- The sample is kept in motion so that it can not freeze on its own.
- The sample is thoroughly mixed so that all parts of the sample have the same temperature.
- The warmth contained in the sample is transported out where it can be dissipated by the cooling mechanism.

When a liquid is colder than its actual freezing point, this state is instable. This state is called "metastable". Even the smallest influences, such as the impact of a hard object on the glass wall, cause freezing to set in. This continues like an avalanche until the released fusion heat increases the temperature of the sample so much that the freezing point is reached and the frozen parts of the sample are in equilibrium with the not yet frozen parts of the sample.

A cryoscope must therefore trigger freezing when the sample is sufficiently colder than its actual freezing point. But what is "sufficiently colder"? The aim here is that so much ice builds up during freezing that there are normal-sized ice crystals all throughout the sample but that the sample is not completely frozen. With milk, it has been proved optimal to trigger the freezing at about -2°C to -3°C .

After triggering freezing, the temperature of the sample climbs because the fusion heat created during freezing is released. It stabilises at a certain value, which is called the plateau. The cooling bath continues to pull warmth out of the sample, and to the same degree that this happens, more parts in the sample freeze and release their fusion heat. Therefore the temperature remains the same – at least as long as there are still liquids parts in the sample. This plateau lasts for a few minutes. The cryoscope determines the freezing point from the temperature measurement values of the plateau. There are rules regulating this.

POSSIBLE SOURCES OF ERROR IN THE MEASUREMENT PROCESS

When determining the freezing point, a certain procedure must be adhered to during measurement, whereby errors can occur at every stage of the procedure.

■ Errors during cooling:

If the heat withdrawn from the sample is too little, the cooling takes too long.

The reason for this is either the cooling bath or the stirring rod. The cooling bath must be at least 6°C and circulate well in order to be able to transfer the heat out of the sample. The stirring rod must stir uniformly with an amplitude of 3.4 mm. When cooling errors occur, the cooling bath temperature must therefore be measured with a thermometer, then the cooling bath circulation is checked with an empty sample tube. Then, it is determined whether the stirring rod can move freely and that it does not strike or grind against anything. Finally, the stirring rod amplitude is tested. There is a special menu in the device for this purpose. The valid reference value is not simply some number on the display; this is only meant to be an indication. The tip of the oscillating stirring rod is observed and adjusted so that the points of regression are only about 3-4 mm apart. Then 2.5 ml of water is poured into a sample tube which is held under the thermistor so that the stirring rod stirs the water. It is determined whether the stirring rod oscillates well in the water.

When everything has been tested and adjusted, a sample measurement with water is carried out and the temperature value in display is observed. The time that the device takes to cool one sample from room temperature (20 °C to 25°C) to -2°C should be almost exactly one minute. If this is the case, it means that the cooling bath and the stirring rod are adjusted correctly. If cooling takes less than 45 seconds, then the cooling bath is too cold or the stirring rod setting is too high. If cooling takes longer than 75 seconds, the cooling bath is too warm or is circulating poorly or the stirring intensity is too low.

If an "error during cooling" occurs after the cooling bath and stirring rod have been tested and determined to be functioning correctly, then the thermistor and the calibration of the instrument must be tested. If the instrument has been incorrectly calibrated, it will not find its temperature scale and therefore cannot measure the temperature correctly.

■ Frozen too early

The state of the sample is instable when it is below its freezing point. It can therefore happen that the sample freezes due to an unintentional influence or on its own before the device triggers freezing. There are many possible reasons for this. If stirring is too strong or if the stirring rod is grinding against something, jolting can occur and trigger freezing. The longer cooling takes the more time the sample has to freeze on its own. Therefore the cooling should be carried out as quickly as possible. If the sample is contaminated, freezing may be triggered.

■ Not frozen:

If the temperature set for supercooling (the "trigger temperature") is reached, the device beats against the glass wall of the sample tube to trigger freezing. The temperature should then start to rise. A criterion for this is a rise in temperature of at least 0.1° C. This is always the case with water or calibration solutions if the stirring rod is set in such a way that it beats hard against the glass wall. This is not always the case with milk. Some milks freeze slowly. Should this error occur rarely with individual milk samples, the milk in question should be heated to approx. 40°C, cooled and measured again. However, if this error occurs often in a certain region, then it is better to lower the trigger temperature so that the samples are supercooled more aggressively, causing them to freeze easier. If this error occurs with calibration solutions, then the calibration of the device is incorrect or cooling bath liquid has leaked into the sample.

■ Plateau not found:

This error can only occur when the "Plateau Search Method" in accordance IDF is used to determine the freezing point. With this method, the temperature value must be within the defined range for a certain time during the plateau. It can so happen that a certain milk sample does not fulfil this criterion. Then a second sample of this milk must be measured. If this error occurs frequently even though the device is otherwise functioning correctly, the error is either with the thermistor or the result of external disturbances.

■ Uncalibrated or defective thermistor:

The instrument tests the current thermistor value when starting a measurement or calibration. Its electrical resistance is known to be a function of the temperature. This electrical resistance is translated with an ADC (analogue digital converter) into a number which is then used by the instrument. If the thermistor has a short circuit or a disruption, its resistance is zero or infinite, both of which conditions are impossible for a properly functioning thermistor. In this case, the thermistor will not start the measurement.

If the temperature which is given from the current thermistor value together with the calibration constant stored in the device is lower than +1°C (which is not possible with a thermistor which is located in a new, i.e. still warm sample), the device will also fail to start the measurement.

IDENTIFYING TECHNICAL DEFECTS

Switching on: the device must show the starting message „CryoStar I (or: CryoStar automatic), Funke Gerber“ on the display when it is switched on.

Possible errors:

- Locking devices on the network connection block
- Locking device on the main conductor board
- Main rectifier. Verify that the voltage of the main condenser is at least 11 V.
- Power transformer
- Error with the main conductor board
- Display or a cable leading to the display is defective

Cooling phase: the device should reach a cooling bath temperature of at least -6°C in a reasonable amount of time. This time depends on the surrounding temperature, but should not be longer than 20 minutes.

Possible errors:

- Air supply is not functioning properly: ventilation slots on the sides of the device are clogged, the inside of the device is contaminated.
- A ventilator has failed.
- Ventilator control system is defective. Verify that the voltage is approx. 24-26 V.
- Cooling block has suffered heating damage and is now defective.
- Cooling block control system is defective. Verify that the Peltier connectors are approx. 6-10 V at full cooling capacity.
- No or poor circulation: when an empty sample tube is immersed (with lid removed) into the measuring site and taken back out, the cooling bath liquid should flow back in within approx. 1 to 2 seconds. **Possible errors:**
 - Cooling bath liquid has become too thick. Change the liquid.
 - Too little cooling liquid, therefore air in the lead: add liquid.
 - Pump is blocked. Switch off device, open lid, carefully turn the pump motor rotor by hand: it should spin without resistance. If this is not the case (contaminants in the pump): rinse pump and lead.
 - Pump control system is defective. Verify that the voltage on the pump motor connectors is approx. 24-26 V.
 - Pump motor is defective: replace motor.
 - Axle between pump motor and pump is defective: remove pump motor, check axle.
 - Device reports the signal “lift error” when starting a measurement. **Possible errors are:**
 - Final position switch on the lift is defective.
 - Cable from measuring head to main conductor board is defective.
 - Conductor board in measuring head is defective.
- Device indicates a much too cold value on the display immediately after starting a measurement, beats the sample tube and reports “not frozen”. This only occurs with old firmware versions. **Causes:**
 - Thermistor is defective. Change thermistor, install newer firmware version.
 - Stirring rod cannot be properly adjusted. **Possible causes:**
 - Stirring rod has been bent during a thermistor change and is touching the thermistor shaft. Bend the stirring rod back into shape and adjust the thermistor so that the stirring rod can oscillate freely.
 - Upper part of the stirring rod has a fatigue fracture: replace stirring rod.
 - Stirring rod was assembled backwards. The magnet in the stirring rod must be orientated in such a way that it is pulled by the current-carrying reel and is not pushed away. Assemble the stirring rod in the correct position.
- Device measures and can be calibrated, but measurement values are scattered. **Possible causes:**
 - Thermistor is defective. Somewhere on the thermistor, microscopic cracks have formed which moisture can now seep through. This causes the electrical properties of the thermistor to become compromised, meaning that the thermistor must be replaced.
 - Impure specimen dishes.
 - Cooling bath liquid has reached the thermistor shaft. A measurement was started without a sample tube. This means that the thermistor was dipped into the cooling bath liquid and the remains of it stuck to the thermistor shaft and have slowly got into the sample.

IDENTIFYING OPERATIONAL ERRORS

Most errors that are made during operation of the device are incorrect calibrations. The calibration of a cryoscope is a precondition for each and every use. For measuring reasons it is necessary to use a thermistor to measure the temperature of a sample. Thermistors have a very strong temperature effect which is necessary for resolutions of more than 1 m°C. Unfortunately, the production-oriented fluctuation range of the resistance values of these components is so large that the temperature zero point (0°C) must usually be determined by a pre-calibration before the device can be calibrated with a new thermistor.

It has to be assumed that the A calibration cannot be successfully executed after a thermistor replacement. The reason for this is that the device first has to reach the set "trigger temperature" and then must recognize a rise in temperature after the glass wall has been hit (as a sign that the freezing has started). This does not happen because the values of the new thermistor result in false temperatures being given when calculated with the calibration constant of the old thermistor. Therefore a so-called pre-calibration is necessary, in which the device ignores the temperatures and follows a purely time-controlled measuring procedure. The calibration constants are subsequently adapted to the characteristics of the new thermistor so that both the A and the B calibration can be successfully carried out.

Unfortunately, it often happens that during the calibration sample tubes filled with calibration solution are taken for something else or that the incorrect menu item is selected.

■ MIX UP: A CALIBRATION INSTEAD OF B CALIBRATION

The entire temperature scale of the device is displaced. When re-measuring the calibration solutions, reversed values and a reversed sign are given.

Example: A cal. with 0.000
A cal. with 0.000
B cal. with -0.557
A cal. with -0.557 (faulty operation)
Re-measuring solution B: results in 0.000
Re-measuring solution A: results in 0.557

■ MIX UP: TAKING THE A SOLUTION INSTEAD OF THE B SOLUTION

At first the A calibration goes as expected. However, when it comes to the B calibration, the device reports the error "uncalibrated" or "thermistor defective" and remains uncalibrated.

■ Defective thermistor

This is a frequently occurring error. There are two possibilities:

1. The thermistor is (was) broken. This can be identified because display constantly shows a negative value that doesn't change.
2. The thermistor bonding is permeable. This can be identified by extremely instable measurement behaviour. The reproducibility is very poor, e.g. there are variations of approx. $\pm 0.1^{\circ}\text{C}$. In both cases the thermistor must be replaced.

■ Stirring rod errors

- ❑ **The stirring rod does not oscillate freely:** it must be able to move freely in the slot provided. It cannot be allowed to touch the thermistor at any place. This must be kept in mind when replacing the thermistor.
- ❑ **The stirring rod amplitude is not high enough:**
The cooling of the sample is not carried out uniformly and takes much longer than 1 minute. With a correctly adjusted stirring rod, the cooling time is almost exactly 1 minute. The stirring rod amplitude must be approx. 3-4 mm. If necessary, the stirring rod must be adjusted accordingly.
- ❑ **The stirring rod amplitude is too high:**
Premature freezing of the sample occurs frequently.

SPECIAL APPLICATIONS/MEASUREMENT OF CREAM

Since the liquid relevant for the freezing point only exhibits 60 % sample volume with a cream of approx. 40 %, it is recommended to increase the sample volume to 3 ml. In addition, the trigger temperature should be set to -3°C, or -3.2°C if the sample repeatedly fails to freeze. It is also possible to marginally increase the impact force of the stirring rod.

Suggested set points

Description	Set points
A calibration	0.000°C or -0.408°C
B calibration	-0.557°C or -0.600°C
Base value	-0.520°C (EU boundary value) Serves solely for calculation of the infiltration water content percentage.
Trigger temperature	-2.00°C (-3.00°C minimum)
Mode	Celsius
Plateau	Plateau search: 0.4 m°C / 22s
Fixed time:	50 s
Maximum:	0.2m°C
Language	free choice
Stirring rod / amplitude	3 - 4 mm
Stirring rod / frequency	Note: Do not change the set value! The values lie between 95 Hz and 104 Hz, depending on the device. Stirring rod/impact force. The impact force should be set to be so powerful that a relatively loud noise is heard when the trigger temperature (e.g. -2°C) is reached. However, it should be seen to that the impact force is not too strong, as this could lead to breakage of the sample tube. The set points lie between approx. 40 % and 50 %.

If the set points are changed, the device must be re-calibrated.

CryoStar[®] automatic

CryoStar I



Important:

The freezing curve is illustrated on the graphic colour display during the measurement. This enables the plateau search in accordance with the prescribed standard (DIN/ISO/IDF/5764) to be ideally tested and reproduced (patent-registered).



Quick and reliable measurement of the freezing point in milk with the CryoStar
Reference measurement in accordance with DIN / ISO / IDF 5764

THE MOST IMPORTANT FEATURES AT A GLANCE:

- **Forward-looking and flexible:** fixed-time measurement, plateau search and maximum search features are available. All parameters relevant to these features can be programmed freely, and, of course, recorded as well. This means that the device can be adjusted to all national and international standards.
- **Easy-to-use:** operation is menu-assisted in the language of your choice. Currently, German, English, French, Greek, Italian, Polish, Portuguese, Spanish, Turkish and Hungarian are available.
- **Efficient:** a new cooling system provides for quick operational readiness even at high surrounding temperatures (up to approx. 32°C).
- **Fast:** up to 40 samples can be measured per hour, depending on the setting.
- **Multifunctional:** the device has a parallel connection (for standard printers) and can be hooked up to a PC with a serial interface. This makes it possible to map the freezing curve on the screen during a measurement and, when necessary, to save it. An efficient zoom function tops off the image. The software needed for this is included in the scope of delivery.
- **User-friendly:** the operation of this device is uncomplicated. The percentage of infiltration water is immediately indicated and printed out. The calibration is executed automatically. All settings and calibrations are permanently saved to non-volatile storage.

Technical specifications:

Connection:	230V/115 V AC (50...60 Hz), 180 VA
Measurement resolution:	0.0001°C (0.1 m°C)
Reproducibility:	± 0.002°C (± 2.0 m°C)
Measuring range:	0.0000°C to -1.5000°C
Sample volume:	2.0 ml to 2.5 ml <i>(recommended value: 2.2 ml)</i>
Sample turnover:	up to 40/h, typically 30/h
Interfaces:	1 x parallel, 1 x serial (RS232)
Cooling time:	approx. 15 min.
Display:	graphic colour display, freezing curve, measurement result [°C], [% infiltration water], date, time, measurement conditions
Protocol printing:	measurement result [°C], [% infiltration water], date, time, measurement conditions

CryoStar I (single sample device)
Automatic cryoscope

Reference method in accordance with ISO/IDF/DIN 5764
This device differs from the "CryoStar_{automatic}" only in
the sample feed system.

Weight: 12.0 kg (net)
Dimensions: 290 x 380 x 190 mm (w x d x h)
With measuring head: 240 mm (h)

7150



CryoStar_{automatic} (multi-sample device)

The measurement procedure of this device is identical
to that of the single sample device "CryoStar 1".
It differs from the "CryoStar 1" only in the sample feed
system. In addition, this device is equipped with a round
magazine for 12 samples. This makes fully automatic
measurement of 12 samples possible with the push of
a button.

Weight: 14.6 kg (net)
Dimensions: 440 x 440 x 200 mm (w x d x h)
With measurement head: 240 mm (h)

7160



Accessories/Expendable items

Thermal printer protocol printer (6 V DC)
for direct connection to the devices
CryoStar (art. no. 7150, 7160) and
LactoStar (art.no. 3510, 3530). Please see art. no. 7157
for compatible thermal paper rolls.

7151

Replacement thermistor,
for CryoStar I and CryoStar_{automatic} (art. no. 7150, 7160)
in accordance with ISO/DIN 5764, PVC, white

7152

Software
for CryoStar (included in the scope of delivery)

7156

Thermal paper roll
for thermal printer art. no. 7151

7157

Connecting cable (12 V DC)
for CryoStar 12 Volt connection

7159

7165 Calibration standard "A"
0.000°C, in 250 ml PE bottle

7166 Calibration standard "B"
-0.557°C, in 250 ml PE bottle



7167 Sample tube
mit Marke with marking at 2.0 ml, 50 pieces

7168 Sample stand
PPH, for 27 sample tubes (art. no. 7167)



7169 Cooling bath liquid
in 500 ml PE bottle



7174 Sampling pipette
adjustable from 1.0 to 5.0 ml

7175 Pipette tips
for art. no. 7174



7186 Calibration standard A
-0.408°C, in 250 ml PE bottle

7187 Calibration standard B
-0.600°C, in 250 ml PE bottle

7188 Confirmation standard C
-0.512°C, in 250 ml PE bottle



Lactometer

easy-to-use hand refractometer for the in-house determination of SNF.

7500

Solubility index mixer

for determining the solubility of milk, cream, whey powder, among others

in accordance with ADPI and DLG regulations, with special motor, glass mixing bowl, stainless steel stirrer, timer and continuous operation switch. See also art. no. 3634

7610 Solubility index mixer

7620 Replacement glass mixing bowl

7621 Replacement stirrer

7622 Replacement drive belt



Reference table

ADPI "Scorched Particle Standards of Dry Milks", 4 stages

7650

Jolting volumeter

Type STAV II for determining the jolting volume of powdered milk.

White plastic casing, high gloss with single-phase AC motor 220 V/50 Hz, jolting mechanism with tension lock for measuring cylinder, five digit electrical pre-selection counter, on/off switch with control lamp, red semi matte control panel. The 250 ml measuring cylinders are standardised by weight and graduation in accordance with DIN 53194

7660

Replacement measuring cylinder

7661 for art. no. 7660



SHORT TIME HEATING DETECTION

determination of alkaline phosphatase

Lactognost original pack

with reference table for 100 samples,
1 spoon

7820

Lactognost refill pack

with reagents I, II and III
for 100 samples

7821

Testing strips Phosphatesmo MI,

pack of 50 strips

7822

Peroxtesmo MI

high temperature heating detection/UHT test
determination of peroxidase

7825

pack of 100 strips

Mastitis detection

LactoStar is used to diagnose a mastitis infection
(see art. no. 3510). In addition, determination by means
of the California Mastitis Test can be done.

California Mastitis Test (CMT) (shalm test)

for quick determination of increased cell content in
milk from which a possible mastitis infection can be
diagnosed

2 test trays with 4 dishes
1 injection flask 250 ml

7920


California Mastitis Test (CMT)

(test liquid)

7930

1 litre

7931

5 litres

Test tube

thick-walled, 100 pieces,

8100 160 x 15 x 16 mm

**Coli tube**

20 x 10 mm, 100 pieces

8120

Durham tube

40 x 8 mm, 100 pieces

8130

Coli tube stand

for 54 samples

stainless steel, sterilisable

8140 150 mm x 100 mm x 205 mm (w x h x d), 600 g

**Sterilizing box for pipettes**

stainless steel

8190 300 x 65 mm (length x thickness)

8191 420 x 65 mm (length x thickness)




8201 **Kapsenberg cap**
various colours




Dilution flask
borosilicate glass 3.3

8290 **250 mL**, with glass rod and silicon stopper,
sterilisable



8291 without accessories
 $l_4 = 190 \text{ mm}$, $l_3 = 20 \text{ mm}$, $l_2 = 27 \text{ mm}$, $l_1 = 143 \text{ mm}$, $d_2 = 20.5 \text{ mm}$, $d_1 = 52 \text{ mm}$



Dilution pipettes

8300 **1.1: 0.1 mL**
 $l = 250 \text{ mm}$, $\varnothing = 5.9 \text{ mm}$

8301 **1.0 + 1.1 mL**,
according to Demeter's method, with 2 markings
 $l = 225 \text{ mm}$, $\varnothing = 6.9 \text{ mm}$

8302 **1.0 + 2.0 + 2.1 + 2.2 mL**,
according to Demeter's method, with 4 markings
 $l = 260 \text{ mm}$, $\varnothing = 6.3 \text{ mm}$

8303 **1.0 + 1.1 + 1.2 mL**,
according to Demeter's method, with 3 markings
 $l = 225 \text{ mm}$, $\varnothing = 7 \text{ mm}$



Petri dish
glass

8310 100 x 20 mm



Petri dishes

plastic (disposable),
sterile packaging

8312 1620 pieces, without vent cam, Ø 55 x 15 mm

8313 480 pieces, with vent cam, Ø 94 x 16 mm

8314 480 pieces, without vent cam, Ø 94 x 16 mm

Sterilizing box

with insert,
stainless steel, for Petri dishes

8320 250 x Ø 120 mm



Wire cages

for sterilisation

8330 100 x 100 x 100 mm

8331 140 x 140 x 140 mm

8332 200 x 200 x 200 mm



Smear needle

rectangular bend

8340 0.59 mm thick

Drigalski spatula

glass

8350 l = 150 mm, triangle height = 30 mm



Inoculation wire

stainless steel

8370 1 m, Ø = 0.46 mm

Burri loop

platinum, calibrated

8380 0.001 ml

8381 0.01 ml

Needle holder

8382 for inoculation wire loop



Slide

for microscope (art. no. 8761, 8762)
half white, cut edges,
50 pieces

8400 76 x 26 mm

Cover glass

for microscope (art. no. 8761, 8762)

8401 18 x 18 mm

8410 Tweezers for slides



8420**Staining stand**

according to Bongert's method

**8430****Staining cuvette**

rectangular

**8440****Wire mesh**

with ceramic centre

8441

without ceramic centre

8450**Tripod** for Bunsen burner**ColonyStar bacteria counter**

easy-to-clean plastic casing, height adjustable illuminated area of 145 mm Ø with direct or indirect glare free lighting, frosted glass and clear glass plate with cm² und 1/9 cm² graduation, electrical contact pin with felt tip pen for marking Petri dishes of up to 145 mm Ø can be used. The supplied reducing insert can be used for dishes with smaller diameters.

220 V/50 Hz, 250 x 230 x 75 mm, 1.7 kg

8500**ColonyStar**

with accessories (art. no. 8501, 8503, 8504, 8505)

8501**Magnifying lens** with sturdy base a. flexible arm**8502****ColonyStar** without accessories**8502-001**Replacement **frosted glass plate****8503**Automatic **contact pin****8504****Felt pen refill** replacement part for art. no. 8503**8505****Clear glass plate** with dark field

Portable bench autoclaves

with screwed-on control thermometer for rapid and efficient vapour sterilisation at 140°C/2.7 bar or 125°C/1.4 bar. Also suitable for autoclaving small amounts of culture media. Special valves can be supplied on request for 115°C/0.7 bar and 121°C/1.1 bar.

A stainless steel instrument board (Ø 215 mm) and a stainless steel tripod are included in the shipment.

220 - 230 Volt, 50 - 60 Hz, 1.6 KW to 1.75 KW, aluminium, polished silk gloss exterior, thermostatic temperature controller, tested safety (GS)

CV-EL 12 L

Volume 12 L, weight 6.1 kg, diameter 24 cm, interior height 24 cm, maximum usable diagonal 32 cm

8541

CV-EL 18 L

Volume 18 L, weight 7.7 kg, diameter 24 cm, internal height 38 cm, usable diagonal 43 cm

8542

8543 Wire basket



Culture cultivating appliance

for cultivation of individual dairy cultures. Stainless steel culture vessels, 5 L with lid and stirrer, PP plastic casing, microprocessor controller, 8 different sizes from 1 x 5 L to 4 x 20 L

8610 1 x 5 L - vessel, 2 x 0.5 L starter culture flask

8611 2 x 5 L - vessel, 2 x 0.5 L starter culture flask

8612 4 x 5 L - vessel, 4 x 0.5 L starter culture flask

8613 1 x 10 L - vessel, 2 x 0.5 L starter culture flask

8614 2 x 10 L - vessel, 2 x 0.5 L starter culture flask

8615 4 x 10 L - vessel, 4 x 0.5 L starter culture flask

8616 2 x 20 L - vessel, 2 x 0.5 L starter culture flask

8617 4 x 20 L - vessel, 4 x 0.5 L starter culture flask

Test tube shaking device

The shaking function is started by pressing down on the test tube support plate.

Shaded pole motor, 45 Watt, 230V
rpm infinitely variable from 0-2800

8650 110 x 100 x 90 mm (w x d x h)



Magnetic stirrer L-71

without heater,
rpm range 50 - 1250
up to 5000 ml capacity
compact aluminium casing

8690 plate diameter: 155 mm



Magnetic stirrer L-81

with heater, heating plate temperature 50-325°C
rpm range 50 - 1250
up to 5000 ml capacity
compact aluminium casing

8691 plate diameter: 145 mm



Stirring rod

8696 25 x 7 mm

8697 30 x 7 mm

8698 80 x 9 mm

Photometer Spekol 1300

Single jet instrument for spectrum and kinetic measurements in the range of 190 - 1100 nm

with numerical display, equipped with printer interface. Easy handling with pre-programmed methods 230 V, 50-60 Hz, 11.5 kg, temperature range: 15 - 35°C

8700 465 x 365 x 175 mm (w x h x d)

Photometer Spekol 1500

Single jet instrument for spectrum and kinetic measurements in the range of 190 - 1100 nm

with high-resolution LDC-VGA screen, equipped with printer interface. Easy handling with pre-programmed methods 230 V, 50-60 Hz, 11.5 kg, temperature range: 15 - 35°C

8701 465 x 365 x 175 mm (w x h x d)



Manual 4x cuvette changer

8702 for 1 cm or for 5 cm, 10 cm cuvettes

8705 Cuvettes

Binocular microscope

with 45° slanted viewer

stable metal casing with coaxial fine and coarse focus with adjustable end stop. Built-in cross table movement L-R 74, V-H 30 mm.

Built-in illumination 6V/20W, power supply 230 V, 50 Hz. Doppel lens Abbe condenser N.A 1.25 with iris diaphragm, pivotable filter holder, height adjustment, glass filters: blue, green.
(Accessories: art. no: 8400, 8401, 8410)

Eyepiece: 0x planar eyepiece

Objectives: achromatic 4x/0.10; 10x/0.25; 40x0.65, 100 x 1.25 oil immersion

8761

Trinocular microscope

in addition to the binocular model with trinocular sliding tube

(accessories: art. no. 8400, 8401, 8410)

8762



Automatic water distillation apparatus

For generating distilled water with a conductivity of under 2.3 $\mu\text{S}/\text{cm}$ at + 20°C.

Apparatus is fabricated completely from stainless steel 1.401. Wall holder and water supply and discharge hoses are included in the scope of the delivery.

Efficient energy consumption due to use of cooling water heated to 80°C.

Destillatmenge:	4 L / h
Storage tank:	4 L
Cooling water consumption:	50 L / h
Power supply:	220 V / 50 Hz; 3.2 kW
Dimensions:	510 x 460 x 230 mm
Weight:	13 kg

8771

Distillate amount:	7 L / h
Storage tank:	7 L
Cooling water consumption:	70 l / h
Power supply:	220 V / 380 V / 50 Hz; 4.8 kW
Dimensions:	670 x 500 x 340 mm
Weight:	19 kg

8772



Water bath

with digital clock up to 999 hours
and temperature rise safety

8786

7 L with gable cover
approx. 11 kg, 240 x 20 x 140 mm

8788

22 l with gable cover
approx. 16 kg, 350 x 290 x 220 mm

THE USE OF REFERENCE MATERIAL IN THE LABORATORY



Dr. Ulrich Leist, DRRR GmbH, reports

Dr. Ulrich Leist studied chemistry and marine life science in Marburg, Stuttgart and Oldenburg, where he got his doctorate in the field of surface science. Afterwards, he was employed as a postdoctoral researcher for a year at Harvard University, Cambridge USA. He gathered additional experience in the field of interlaboratory tests/reference material during his four year employment at Muva Kempten. Since 2007 he has been the executive director of the Deutsches Referenzbüro für Lebensmittelringversuche und Referenzmaterialien GmbH (DRRR GmbH) (German Reference Office for Foodstuff Interlaboratory Tests and Reference Material, Ltd.).

Fundamentals for evaluating laboratory results of the main parameters in dairy farming

The use of reference material in the laboratory serves to assure quality. On the one hand, laboratory personnel can be trained, methods can be developed, checked and optimized, and measuring devices can be tested for their operational capability, accuracy and precision. Of particular importance when doing this is the calibration of indirect measurement equipment, e.g. IR spectrometers, with which the measurement signal is first related to the reference parameter, for example to a measurement parameter such as fat.

To ensure the optimal use of reference material, the fundamental terms should be briefly defined:

- **Accuracy:** degree of the total error of an analysis and thereby an umbrella term for correctness and precision.
A result is accurate when it is free of incidental and systematic mistakes.
- **Correctness:** degree of deviation from the measurement value (or, the mean of many measurement values) to the correct (actual) value due to a systematic mistake (*also: bias for the amplitude of a systematic mistake*).
- **Precision:** precision indicates how widely the analysed values are scattered due to incidental mistakes. Precision is statistically described by the standard deviation or the confidence interval.
- **Reproducibility (repeatability limit) r:**
The absolute difference between two single measurement values that can be expected from the same material, the same methods, the same person, the same instrument, the same laboratory and the same time frame with a probability of 95 %.
- **Comparability (comparability limit) R:**
The absolute difference between two single measurement values that can be expected from the same material, the same methods, different people, different instruments, different laboratories and a larger time frame with a probability of 95 %.

Precision data for the methods is of particular importance as this makes it possible for laboratories to evaluate whether they are proficient in a method and whether measurement results of different laboratories are comparable. This is crucial in the case of reference methods because they are the accepted foundation on which products like foodstuffs can be judged. Precision data is documented in various standards and official regulations.

Precision data r and R from:

- DIN/EN/ISO
- IDF
- § 64 LFGB (previously: § 35 LMBG)
- VDLUFA

Precision data for milk

Parameter	Method	r	R	s _R	CRD	Range of application
Fat	Roese Gottlieb	0.02 % 0.02 % 0.01 %	0.04 % 0.03 % 0.025 %	0.014 % 0.011 % 0.009 %	0.026 % 0.019 % 0.017 %	3.5 % fat 1.5 % fat (0.5 to 2 %) skim milk <0.5 % fat
Dry matter	102°C, sea sand	0.10 %	0.20 %	0.071 %	0.132 %	
Protein	Kjeldahl	0.04 %	0.10 %	0.035 %	0.068 %	
Lactose	enzymatic determination	Value x 0.05	Value x 0.06	$\frac{R}{2.83}$		
Freezing point	cryoscopy	0.004°C	0.006°C	0.002°C	0.004°C	

Precision data for powdered milk

Parameter	Method	r	R	s _R	CRD	Range of application
Fat	Roese Gottlieb	0.2 % 0.15 % 0.1 %	0.3 % 0.25 % 0.2 %	0.106 % 0.088 % 0.071 %	0.187 % 0.160 % 0.132 %	VMP, powdered cream Partially removed powder Skim milk powder
Dry matter	102°C, sea sand	0.2 %	0.4 %	0.141 %	0.265 %	
Protein	Kjeldahl	0.3 %	0.8 %	0.283 %	0.545 %	
Lactose	enzymatic determination	Value x 0.05	Value x 0.06	$\frac{R}{2.83}$		

Precision data for processed cheese

Parameter	Method	r	R	s _R	CRD	Range of application
Fat	SBR	0.1 % 0.2 %	0.4 % 0.6 %	0.141 % 0.212 %	0.278 % 0.412 %	10 % abs. fat 25 % abs. fat
Dry matter	102°C, sea sand	0.3 %	0.5 %	0.177 %	0.320 %	
Protein	Kjeldahl	0.19 %	0.38 %	0.134 %	0.251 %	
Lactose	enzymatic determination	Value x 0.05	Value x 0.06	$\frac{R}{2.83}$		

QUALITY FROM THE VERY BEGINNING

The use of modern analytical reference systems for the processing of milk is characterized by challenging analytical and statistical demands.

Milk processing is accompanied by a series of measures which assure quality. These measures of course include the analysis of milk beginning at the moment the milk is delivered to the milk processing company, namely dairies. For the analysis of chemical quality parameters of milk such as protein, fat, lactose, dry matter and freezing point, infrared spectroscopy methods as well as thermo-analytical methods (LactoStar) are used extensively in the processing of milk. When doing so, the use of modern IR spectrometers makes it possible to provide test results for the above-mentioned testing parameters within just a few seconds. The speed advantage over the reference testing methods such as Roesse Gottlieb's method for determining fat content (test duration approx. 8 hours) or Kjeldahl's method for determining protein content (test duration approx. 8 hours) is enormous. This speed advantage enables rapid response to changes in the constitution of milk as well as with milk supply and the intermediate and end products and makes it possible to adjust the production accordingly. This means that e.g. the fat or protein content for the respective product can be held constant throughout the production time. The IR and thermal analysis not only lend themselves to the control of raw milk but also to all intermediate and end products.

The only disadvantage of the IR spectroscopic methods and thermo-analytical methods is the fact that these are indirect methods. That means that the analytical instruments have to be calibrated.

Calibration

For calibration, a concentration value must be related to the measurement signal of the analytical instrument.

This fundamental calibration is usually included in the scope of delivery of the instrument or is configured with the help of the instrument manufacturer.

The basic calibration is ultimately a relation of the physical measurement scope to a substance. The regular calibration related to a concrete product is usually carried out by the user himself. Here, the change in concentration of a substance is ultimately correlated to the change in the strength of the signal. Conventionally, to do this a sample is measured with the analytic instrument whose substance is determined at the same time through tests with rapid methods or reference methods. The test results obtained are then allocated to the analytical instrument during calibration. At the same time, this means that the uncertainty of measurement of the test results must be factored in to the instrument. The precision reached with the calibration can hence only lie within the boundaries of the method comparability. This means that the theoretically possible precision is entered into the analytical instrument. In addition, the measurement speed advantage is partially quantified by the increase in measurement certainty. In order to cancel out this disadvantage, the number of calibration sample tests with rapid or reference methods can be increased, which is however very costly. Since it is necessary to regularly calibrate the analytical instrument, each and every increase in the sample number for tests in the context of calibration value assignment leads to an increase in test complexity and expense.

However, consideration of measurement uncertainty can not be foregone. When fulfilling official regulations regarding foodstuffs, e.g. pasteurized milk, these uncertainties during calibration must be considered. If the target values for a product are not fulfilled, it can lead to penalties from the customer or to violation of food labelling regulations.

Modern analytical instruments boast precise measuring technology, but reliable and accurate calibration is essential to complement it.

The means of selecting accurate and precise calibration lies in the use of verified reference systems.

A reference system is ultimately dependent on the reference material which has been confirmed by interlaboratory tests. Particularly high demands are placed on quality assurance with a reference system like that of the DRRR corporation, which has earned the status of a leading proficiency level.

The reference values are determined by an interlaboratory test.

- Only reference methods are used to test reference methods
- The reference laboratories fulfil the demands of standards DIN EN ISO/IEC 17025
- The reference laboratories are under constant supervision, through which they regularly demonstrate their above-average competence by successful participation in respective interlaboratory tests
- The determination of reference values is carried out using extensive modern statistical methods in accordance with the current status of science and technology
- Materials are manufactured without preservatives, using actual shock frost procedures.

The use of reference materials yields the following advantages:

- The materials are related to the reference methods. Thus, the calibration also refers to reference methods. Being that the laboratories are supervised, the return to the reference method is largely assured.
- The material uncertainty corresponds to the comparability of the reference method.
- The use of reference materials makes the simultaneous testing of self-produced calibration samples superfluous. This means lower costs.
- The use of calibration materials assures high linearity, precision and accuracy.
- The materials can be used at any time. Thus, flexibility is increased. In connection with the rapid analysis time with an IR device, a considerable speed advantage is won over classic calibration procedures in dairy laboratories.

CALIBRATION PROCEDURE

There are in essence two calibration procedures. First, the multiple point calibration procedure and second, the one point calibration procedure. A fundamental calibration is required for both procedures.


One point calibration

In this context, the calibration usually only has to do with a bias adjustment. This is by all means acceptable if the calibration itself can be assumed to be stable. If this is not the case, a deviation in the measurement value to the expected reference value can only indicate a general deviation. An adjustment to the instrument setting in the direction of the reference value can even in an extreme lead to a degradation of the calibration state.


Multiple point calibration

Various calibration samples with different concentrations of the target parameters are measured. The various concentrations of the respective substance (parameter) are set in relation to different measurement signal strengths. In doing so, for each calibration sample measured the reference value of the target parameter is related to the measurement signal. Within the area of concentration of the various calibration samples, a mathematical interrelation regarding the calibration slope between the sample concentration and measuring value is produced on the analytical instrument.


Of particular importance with the multiple point calibration is the “rear values”, meaning the values with higher concentration. As can be seen in the following figure, the “rear” value has considerable influence on the calibration slope. If the calibration slope should be steepened, it is advisable to set additional calibration points in the high (rear) area of concentration. The goods of the calibration can be read off on the correlations coefficient, among other things. This should by all means be higher than 0.9. The correlation coefficient indicates how probable it is that the calibration points actually match the calibration slope. The maximum correlations coefficient that can be reached is 1 (=100 %).




Normal 6 point calibration, with a correlation coefficient near 1.



The 6th measurement value (rear value) is low. The correlation coefficient is near 0.9.




6 point calibration, the 6th measurement value is high, correlation coefficient near 1.



Range of calibration slope, with uncertainties in “rear” value area.

In addition to the use of reference material, interlaboratory tests are applied to assure quality. However, the focus of these is often set exclusively on the z-score. For this reason, it will be briefly explained in the following section.

THE Z-SCORE



With the midpoint and the standard deviation the z-score can be calculated for any laboratory using the following equation [2].

$$z - score = \frac{x_i - m}{s}$$

The respective laboratory measurement value (usually the midpoint of the repeat determination) x_i is set in the equation above. Then the midpoint m and the standard deviation s of the entire data set are entered into the equation. Thus, the distance of the laboratory value to the midpoint is calculated in units of standard deviation. A laboratory which has a z-score of exactly 2 has a distance to the midpoint of exactly 2 standard deviations. That means that the laboratory is just barely part of the 95.45 % of the values that are expected around the midpoint. In the area between 2 and 3 standard deviations lies the remaining 5 % of the values. A z-score of 3 or larger means that there is only a probability of 0.027 % of belonging to the data set observed. The z-score is assessed accordingly:

$z < 2$	data credible
$2 < z < 3$	3 data questionable
$z > 3$	data not credible

In any case it can be wise for the interlaboratory test participant to select the data sets of a interlaboratory test with which he wants to use for comparison, for example because it uses the same methods, or features the competitor or his customer. He can calculate his own z-score according to his test question with equation 3, which has the necessary informative value according to the test question.


Reference material

You will find the most important reference material for chemical milk analysis with article numbers 3517, 3518, 3519, 3521 (page 45)

LABORATORY GLASSWARE

Beaker

short design, borosilicate glass,
with graduation and spout



8800 50 ml d = 38.7 mm, l = 60 mm

8801 100 ml d = 47 mm, l = 70 mm

8802 250 ml d = 67 mm, l = 95 mm

8803 400 ml d = 76.2 mm, l = 110 mm

8804 600 ml d = 86.6 mm, l = 125 mm

8805 800 ml d = 94 mm, l = 135 mm

8806 1000 ml d = 102 mm, l = 145 mm

tall design, borosilicate glass,
with graduation and spout

8808 50 ml d = 34.6 mm, l = 71 mm

8809 100 ml d = 44.5 mm, l = 80 mm

8810 250 ml d = 57 mm, l = 122 mm

8811 400 ml d = 67 mm, l = 129 mm

8812 600 ml d = 77.9 mm, l = 148 mm


8813 800 ml d = 84 mm, l = 175 mm

8814 1000 ml d = 92.8 mm, l = 180 mm

8815 2000 ml d = 114 mm, l = 240 mm

Erlenmeyer flasks

narrow necked, borosilicate glass
with graduation, DIN 12380



8817	50 ml	$d_2 = 19.4 \text{ mm}$, $l = 87 \text{ mm}$, $d_1 = 51 \text{ mm}$
8818	100 ml	$d_2 = 17.9 \text{ mm}$, $l = 108 \text{ mm}$, $d_1 = 63.5 \text{ mm}$
8819	200 ml	$d_2 = 31.1 \text{ mm}$, $l = 135 \text{ mm}$, $d_1 = 78.7 \text{ mm}$
8820	250 ml	$d_2 = 32 \text{ mm}$, $l = 146 \text{ mm}$, $d_1 = 83 \text{ mm}$
8821	300 ml	$d_2 = 31.5 \text{ mm}$, $l = 165 \text{ mm}$, $d_1 = 86 \text{ mm}$
8822	500 ml	$d_2 = 32.3 \text{ mm}$, $l = 180 \text{ mm}$, $d_1 = 104.5 \text{ mm}$
8823	1000 ml	$d_2 = 38.9 \text{ mm}$, $l = 225 \text{ mm}$, $d_1 = 130.3 \text{ mm}$
8824	2000 ml	$d_2 = 46.6 \text{ mm}$, $l = 285 \text{ mm}$, $d_1 = 165 \text{ mm}$


wide necked, borosilicate glass
with graduation, DIN 12385

8826	50 ml	$d_2 = 31.1 \text{ mm}$, $l = 86 \text{ mm}$, $d_1 = 51.4 \text{ mm}$
8827	100 ml	$d_2 = 31.7 \text{ mm}$, $l = 107 \text{ mm}$, $d_1 = 63.5 \text{ mm}$
8828	200 ml	$d_2 = 45.7 \text{ mm}$, $l = 140 \text{ mm}$, $d_1 = 78 \text{ mm}$
8829	250 ml	$d_2 = 47 \text{ mm}$, $l = 140 \text{ mm}$, $d_1 = 84.7 \text{ mm}$
8830	300 ml	$d_2 = 47.6 \text{ mm}$, $l = 154 \text{ mm}$, $d_1 = 87 \text{ mm}$
8831	500 ml	$d_2 = 46.8 \text{ mm}$, $l = 175 \text{ mm}$, $d_1 = 105 \text{ mm}$
8832	1000 ml	$d_2 = 47.8 \text{ mm}$, $l = 215 \text{ mm}$, $d_1 = 132 \text{ mm}$
8833	2000 ml	$d_2 = 64.8 \text{ mm}$, $l = 280 \text{ mm}$, $d_1 = 150 \text{ mm}$




Measuring cylinder

tall design, glass,
with spout



8850	50 ml	1/1 ml, $d_2 = 22.4 \text{ mm}$, $d_1 = 65 \text{ mm}$, $l_1 = 195 \text{ mm}$
8851	100 ml	1/1 ml, $d_2 = 27.5 \text{ mm}$, $d_1 = 76 \text{ mm}$, $l_1 = 245 \text{ mm}$
8852	250 ml	2/1 ml, $d_2 = 36.5 \text{ mm}$, $d_1 = 96 \text{ mm}$, $l_1 = 320 \text{ mm}$
8853	500 ml	5/1 ml, $d_2 = 47 \text{ mm}$, $d_1 = 114 \text{ mm}$, $l_1 = 380 \text{ mm}$
8854	1000 ml	10/1 ml, $d_2 = 61 \text{ mm}$, $d_1 = 145 \text{ mm}$, $l_1 = 465 \text{ mm}$



Measuring cylinder

tall design, polypropylene,
blue graduation

8855	50 ml	1/1 ml, $d_2 = 23$ mm, $d_1 = 68$ mm, $l_1 = 200$ mm
8856	100 ml	1/1 ml, $d_2 = 28$ mm, $d_1 = 88$ mm, $l_1 = 260$ mm
8857	250 ml	2/1 ml, $d_2 = 42$ mm, $d_1 = 101$ mm, $l_1 = 310$ mm
8858	500 ml	5/1 ml, $d_2 = 61$ mm, $d_1 = 95$ mm, $l_1 = 350$ mm
8859	1000 ml	10/1 ml, $d_2 = 70.5$ mm, $d_1 = 135$ mm, $l_1 = 415$ mm
8860	2000 ml	20/1 ml, $d_2 = 81$ mm, $d_1 = 160$ mm, $l_1 = 490$ mm

Mixing cylinder



AR glass, round base,
with NS PE stopper

8862	100 ml 1/1	$d_2 = 22.4$ mm, $d_1 = 58$ mm, $l_1 = 280$ mm
8863	250 ml 2/1	$d_2 = 27.7$ mm, $d_1 = 85$ mm, $l_1 = 340$ mm

Measuring flask

with stopper,
borosilicate glass, with ring markings
DIN 12664, adjusted to "In"



8870	25 ml	$d_2 = 6.5$ mm, $d_1 = 37$ mm, $l_1 = 38$ mm, $l_2 = 73$ mm, $l_3 = 111$ mm
8871	50 ml	$d_2 = 12$ mm, $d_1 = 48$ mm, $l_1 = 45$ mm, $l_2 = 92$ mm, $l_3 = 137$ mm
8872	100 ml	$d_2 = 11.09$ mm, $d_1 = 60$ mm, $l_1 = 63$ mm, $l_2 = 111$ mm, $l_3 = 174$ mm
8873	250 ml	$d_2 = 12.9$ mm, $d_1 = 78$ mm, $l_1 = 85$ mm, $l_2 = 130$ mm, $l_3 = 215$ mm
8874	500 ml	$d_2 = 17.3$ mm, $d_1 = 100$ mm, $l_1 = 110$ mm, $l_2 = 150$ mm, $l_3 = 260$ mm
8875	1000 ml	$d_2 = 22$ mm, $d_1 = 126$ mm, $l_1 = 140$ mm, $l_2 = 165$ mm, $l_3 = 305$ mm



Glass funnel

AR glass, smooth,
slanted end,
with short stem, DIN 12445

8876	$d_2 = 55$ mm, $d_1 = 10$ mm, $l_3 = 90$ mm, $l_2 = 40$ mm, $l_1 = 50$ mm
8877	$d_2 = 100$ mm, $d_1 = 10.2$ mm, $l_3 = 180$ mm, $l_2 = 80$ mm, $l_1 = 100$ mm
8878	$d_2 = 150$ mm, $d_1 = 16$ mm, $l_3 = 275$ mm, $l_2 = 130$ mm, $l_1 = 145$ mm
8879	$d_2 = 200$ mm, $d_1 = 20.3$ mm, $l_3 = 330$ mm, $l_2 = 165$ mm, $l_1 = 165$ mm



Measuring pipettes

colour code, AR glass



8882	1 ml, 1/100	$l_4 = 360 \text{ mm}$, $d_1 = 5 \text{ mm}$
8883	2 ml, 1/50	$l_4 = 360 \text{ mm}$, $d_1 = 5.9 \text{ mm}$
8884	5 ml, 1/10	$l_4 = 360 \text{ mm}$, $d_1 = 7.5 \text{ mm}$
8885	10 ml, 1/10	$l_4 = 360 \text{ mm}$, $d_1 = 9.9 \text{ mm}$
8886	25 ml, 1/10	$l_4 = 400 \text{ mm}$, $d_1 = 14 \text{ mm}$
8887	50 ml, 1/5	$l_4 = 455 \text{ mm}$, $d_1 = 16 \text{ mm}$



Volumetric pipette

color code, AR glass

8888	1 ml	$l_4 = 325 \text{ mm}$, $l_3 = 135 \text{ mm}$, $l_2 = 35 \text{ mm}$ $l_1 = 155 \text{ mm}$, $d_1 = 4 \text{ mm}$, $d_2 = 6 \text{ mm}$
8889	2 ml	$l_4 = 340 \text{ mm}$, $l_3 = 145 \text{ mm}$, $l_2 = 40 \text{ mm}$ $l_1 = 155 \text{ mm}$, $d_1 = 5 \text{ mm}$, $d_2 = 7 \text{ mm}$
8890	5 ml	$l_4 = 380 \text{ mm}$, $l_3 = 155 \text{ mm}$, $l_2 = 55 \text{ mm}$ $l_1 = 170 \text{ mm}$, $d_1 = 6 \text{ mm}$, $d_2 = 10 \text{ mm}$
8891	10 ml	$l_4 = 450 \text{ mm}$, $l_3 = 200 \text{ mm}$, $l_2 = 70 \text{ mm}$ $l_1 = 180 \text{ mm}$, $d_1 = 7 \text{ mm}$, $d_2 = 12 \text{ mm}$
8892	20 ml	$l_4 = 520 \text{ mm}$, $l_3 = 250 \text{ mm}$, $l_2 = 90 \text{ mm}$ $l_1 = 180 \text{ mm}$, $d_1 = 8 \text{ mm}$, $d_2 = 16 \text{ mm}$
8893	25 ml	$l_4 = 530 \text{ mm}$, $l_3 = 230 \text{ mm}$, $l_2 = 105 \text{ mm}$ $l_1 = 195 \text{ mm}$, $d_1 = 10 \text{ mm}$, $d_2 = 17 \text{ mm}$
8894	50 ml	$l_4 = 550 \text{ mm}$, $l_3 = 245 \text{ mm}$, $l_2 = 120 \text{ mm}$ $l_1 = 185 \text{ mm}$, $d_1 = 7 \text{ mm}$, $d_2 = 26 \text{ mm}$
8895	100 ml	$l_4 = 575 \text{ mm}$, $l_3 = 240 \text{ mm}$, $l_2 = 135 \text{ mm}$ $l_1 = 200 \text{ mm}$, $d_1 = 8 \text{ mm}$, $d_2 = 36 \text{ mm}$





Pipette helper

8920 for pipettes up to 25 ml



Laboratory bottles

borosilicate glass, with ISO thread, graduated,
with PPN screw cap
and PPN pouring ring (blue)



8970 100 ml $l_4 = 105$ mm, $l_3 = 20$ mm, $l_2 = 20$ mm
 $l_1 = 65$ mm, $d_1 = 55$ mm, $d_2 = 31$ mm

8971 250 ml $l_4 = 140$ mm, $l_3 = 25$ mm, $l_2 = 25$ mm
 $l_1 = 90$ mm, $d_1 = 70$ mm, $d_2 = 29.5$ mm

8972 500 ml $l_4 = 180$ mm, $l_3 = 28$ mm, $l_2 = 40$ mm
 $l_1 = 112$ mm, $d_1 = 84.7$ mm, $d_2 = 29.5$ mm

8973 1000 ml $l_4 = 230$ mm, $l_3 = 28$ mm, $l_2 = 48$ mm
 $l_1 = 154$ mm, $d_1 = 100$ mm, $d_2 = 29.5$ mm

8974 2000 ml $l_4 = 270$ mm, $l_3 = 27$ mm, $l_2 = 75$ mm
 $l_1 = 168$ mm, $d_1 = 136$ mm, $d_2 = 29.5$ mm

Wide necked reagent bottles

AR glass, white with standard polish and stopper

8980 50 ml NS 24/20 $l_4 = 87$ mm, $l_3 = 17$ mm, $l_2 = 17$ mm
 $l_1 = 53$ mm, $d_2 = 14$ mm, $d_1 = 45$ mm

8981 100 ml NS 29/22 $l_4 = 96$ mm, $l_3 = 24.5$ mm, $l_2 = 8.4$ mm
 $l_1 = 63.1$ mm, $d_2 = 28$ mm, $d_1 = 53$ mm

8982 250 ml NS 34/35 $l_4 = 142$ mm, $l_3 = 28$ mm, $l_2 = 28$ mm
 $l_1 = 86$ mm, $d_2 = 34$ mm, $d_1 = 142$ mm

8983 500 ml NS 45/40 $l_4 = 167$ mm, $l_3 = 31$ mm, $l_2 = 26$ mm
 $l_1 = 110$ mm, $d_2 = 43.8$ mm, $d_1 = 87$ mm

8984 1000 ml NS 60/46 $l_4 = 200$ mm, $l_3 = 45$ mm, $l_2 = 30$ mm
 $l_1 = 125$ mm, $d_2 = 58$ mm, $d_1 = 109$ mm

8985 2000 ml NS 60/46 $l_4 = 255$ mm, $l_3 = 50$ mm, $l_2 = 41$ mm
 $l_1 = 164$ mm, $d_2 = 58$ mm, $d_1 = 134$ mm



Narrow necked reagent bottles

AR glass, white with standard polish and stopper

8990 50 ml NS 14/15 $l_4 = 77$ mm, $l_3 = 15$ mm, $l_2 = 12$ mm
 $l_1 = 50$ mm, $d_2 = 13$ mm, $d_1 = 42$ mm

8991 100 ml NS 14/15 $l_4 = 105$ mm, $l_3 = 25$ mm, $l_2 = 7$ mm
 $l_1 = 60$ mm, $d_2 = 13$ mm, $d_1 = 52$ mm

8992 250 ml NS 19/26 $l_4 = 135$ mm, $l_3 = 25$ mm, $l_2 = 30$ mm
 $l_1 = 80$ mm, $d_2 = 17.6$ mm, $d_1 = 71$ mm

8993 500 ml NS 24/20 $l_4 = 165$ mm, $l_3 = 47$ mm, $l_2 = 35$ mm
 $l_1 = 100$ mm, $d_2 = 22$ mm, $d_1 = 87$ mm

8994 1000 ml NS 29/22 $l_4 = 205$ mm, $l_3 = 35$ mm, $l_2 = 42$ mm
 $l_1 = 128$ mm, $d_2 = 28$ mm, $d_1 = 108$ mm

8995 2000 ml NS 29/32 $l_4 = 265$ mm, $l_3 = 35$ mm, $l_2 = 70$ mm
 $l_1 = 160$ mm, $d_2 = 29$ mm, $d_1 = 130$ mm



Culture tubes

DURAN glass, straight rim

9050 16 x 160 mm, 100 pieces

Culture tubes

with ISO thread and screw cap
AR glass, sterilisable

9054 16 x 100 mm, 100 pieces

9056 16 x 160 mm, 100 pieces

Test tubes

DURAN glass

9080 without rim, 16 x 160 mm, 100 pieces

9081 with rim, 16 x 160 mm, 100 pieces

Test tube brush

with wool head

9090 length: 230 mm



Weighing dishes

short design, with knob lid

9120 35 x 30 mm

9121 50 x 30 mm



Digital burette type μ 10

without bottle

certified conformity up to 100 mL,
smallest adjustment interval 10 μ L.

bottle: see art. no. 8973

9190



Desiccator, glass, type Novus,
flat flange with knob lid, 250 mm,

9201

Desiccator plate, porcelain

9211

Wash bottles

polyethylene

100 ml

$d_1 = 44.5$ mm, $d_2 = 12$ mm, $l_4 = 105$ mm
 $l_3 = 15$ mm, $l_2 = 27$ mm, $l_1 = 63$ mm

9230

250 ml

$d_1 = 59$ mm, $d_2 = 19.5$ mm, $l_4 = 139$ mm
 $l_3 = 15$ mm, $l_2 = 39$ mm, $l_1 = 85$ mm

9231

500 ml


$d_1 = 74$ mm, $d_2 = 18$ mm, $l_4 = 175$ mm
 $l_3 = 15$ mm, $l_2 = 45$ mm, $l_1 = 115$ mm

9232

1000 ml


$d_1 = 94$ mm, $d_2 = 21.5$ mm, $l_4 = 220$ mm
 $l_3 = 26$ mm, $l_2 = 49$ mm, $l_1 = 145$ mm

9233



Funnels

polyethylene



9235 $d_2 = 40 \text{ mm}$, $d_1 = 9.5 \text{ mm}$, $l_3 = 63 \text{ mm}$, $l_1 = 33 \text{ mm}$, $l_2 = 30 \text{ mm}$

9236 $d_2 = 70 \text{ mm}$, $d_1 = 11.7 \text{ mm}$, $l_3 = 109 \text{ mm}$, $l_1 = 55 \text{ mm}$, $l_2 = 54 \text{ mm}$

9237 $d_2 = 100 \text{ mm}$, $d_1 = 13.8 \text{ mm}$, $l_3 = 155 \text{ mm}$, $l_1 = 80 \text{ mm}$, $l_2 = 75 \text{ mm}$

9238 $d_2 = 120 \text{ mm}$, $d_1 = 15.3 \text{ mm}$, $l_3 = 175 \text{ mm}$, $l_1 = 85 \text{ mm}$, $l_2 = 90 \text{ mm}$

9239 $d_2 = 140 \text{ mm}$, $d_1 = 16.7 \text{ mm}$, $l_3 = 170 \text{ mm}$, $l_1 = 65 \text{ mm}$, $l_2 = 105 \text{ mm}$

Test tube racks

PP plastic, for glass 160 x 16 mm,
sterilisable up to 121°C

9255 12 samples

9256 24 samples



Test tube rack

9257 36 samples, wire, plastic coated

9300 Laboratory lift



Lyphan strips
in plastic screw jar

9360	pH 1 – 11
9361	pH 3.9 – 6.9
9362	pH 4.9 – 7.9
9363	pH 6.9 – 9.9
9364	pH 0 – 14



Indicator paper
for degree of freshness of milk, Duplex

9365	pH 7.9 – 11, 100 pieces
-------------	-------------------------



Burette stand

9400	Plate stand, 210 x 130 x 750 mm
9401	Tripod stand, 210 x 130 x 750 mm



9405 Double socket

9406 Double socket rotatable

Stand clamp without socket

9407 25 mm

9408 60 mm



9409 Stand ring with socket, 160 mm

Burette clamp with socket

9410 single

9411 double

9440 Laboratory clock 0 - 60 min.

Laboratory vacuum pump/compressor
electrical, can be used as a vacuum or pressure pump
max. output 16 L/min.,
max. operating pressure 3.5 bar

9470



Proportioning devices (digital)

for aggressive acids and bases,
without bottle

9484 1 – 10 ml: 0.05 ml,
with thread adapter: A25, A28, A32, A38, A40

9485 2.5 – 25 ml: 0.1 ml,
with thread adapter: A32, A38, A40



Variable proportioning devices

for aggressive acids and bases,
without bottle

9487 1 – 10 ml: 0.2 ml,
with thread adapter: A25, A28, A32, A38, A40

9488 2.5 – 25 ml: 0.5 ml,
with thread adapter: A32, A38, A40



Replacement parts for proportioning devices

Adapter external thread

32 mm for bottle thread A 25 mm
32 mm for bottle thread A 28 mm
45 mm for bottle thread A 32 mm
45 mm for bottle thread A 38 mm
32 mm for bottle thread S 40 mm
45 mm for bottle thread S 40 mm

9489

Microlitre pipettes

variable volume adjustment, with disposable tips

9495 10 – 100 µl

9498 100 – 1000 µl



Pipette tips

9510 1 – 200 µl (yellow), 1000 pieces

9511 50 – 1000 µl (blue), 1000 pieces

ALPHABETICAL INDEX

Description	Art. no.	Page	Description	Art. no.	Page
Accessories for CryoStar [7150]	7151-7188	98-99	Butyrometer tubes, open	3766-0	54
Accessories for humidity measuring device					
MLB 50 [5670]	5671,5672,5673,5674	76	Calibration standard A=0.000°C	7165	99
Accessories for LactoFlash [3530]	7151,3516,3563	45,98	Calibration standard A=0.408°C	7186	99
Accessories for LactoStar [3530]	7151,3511,3516,3563	76	Calibration standard B=0.0557°C	7166	99
Accessories for RD-8 [5700]	5701,5702,5703,5704	77	Calibration standard B=0.600°C	7187	99
Accessories for SuperVario	3631-3633,3685-3687	46,52	Canister	3511	44
Accessories for WB 438 [3707,3708]			Cap for milk sample bottle 80 ml	3043	10
3717-3718,3727,3737,3747,3754,3766		54	Cap for reference drier specimen receptacle	5703	77
Acidity tester	4705	69	Centrifuge NovaSafety	3670	47
Adapter for proportioning devices	9489	130	Centrifuge SuperVario	3680	48-52
Alcoholometer	6710,6711	84	Centrifuge tube according to Friese's method	3638	47
Alkaline phosphatase	7820,7821,7822	101	Centrifuge, safety centrifuge	3680-L	52
Aluminium foil	5470	73	Chamber for thin layer chromatography	5607	74
Aluminium specimen dish			Cheese beaker, perforated	3321,3321-001	32
for humidity measuring device	5671,5674	76	Cheese butyrometer according to Van Gulik's method	3230	29
Attachment for LactoFlash hose pump	3530-023A	45	Cheese trier	3120,3121,3122,3124	10,11
Attachment for LactoStar hose pump	3510-023A	44	Circular glass fibre filter for humidity measuring device	5672	76
Autoclave (portable bench autoclave)	8541,8542	107	Cleaning agent for LactoStar, LactoFlash	3563	44,45
Automatic tilt measure Superior	3420,3421	34	Cleaning agent for single rod electrode	4420	67
			Cleaning brush for butyrometer body	3324	33
Babcock bottles	3254,3256,3258	30	Cleaning brush for butyrometer neck	3325	33
Babcock bucket	3632	46	Cleaning brush for milk sample bottle	3080	10
Bacteria counter Colony Star	8500/8502	106	Cleaning brush for pipettes	3470	35
Bag clasps for BagMixer	3143	11	Clear glass plate for ColonyStar	8505	106
BagMixer	3139,314	11	CMT test liquid	7930/7931	101
Battery/pocket pH meter	4315,4317, 4450	65	Coli tube	8120	102
Beakers	8800-8815	120	Coli tube stand for 54 samples	8140	102
Buckets for ADPI tubes	3633	46	ColonyStar bacteria counter	8500/8502	106
Buffer solutions	4390,4391,4392,4460,4461,4462	66,67	Colour disc for resazurine tests	5161	72
Bunsen burner	5550	73	Confirmation standard C= -0.512°C	7188	99
Burette clamp with socket	9410,9411	129	Connecting cable (12 V DC) for CryoStar	7159	98
Burette stand	9400,9401	128	Contact pin for ColonyStar	8503	106
Burette, digital	9190	126	Contact pin for patent seal FIBU	3270	31
Burri loop	8380,8381	105	Contact pin for patent seal GERBAL	3271	31
Butter beaker with two holes	3323	32	Contact pin for patent seal NOVO	3272	31
Butter butyrometer according to Roeder's method	3220	29	Control thermometer	7070,7071	87
Butter cutter	5605	74	Cooling bath liquid	7169	99
Butter melting beaker	5400,5401	72	Cooling chamber thermometer	7060	87
Butter testing spoon	5450	72	Cover glass for microscope	8401	105
Butter trier	3130,3131	11	Cream and ice cream butyrometer	3189,319	27
Buttermilk tester	6650	83	Cream beaker, unperforated for butyrometers	3320	32
Butyrometer bucket for SuperVarioN			Cream butyrometer according to Koehler's method		
3631,3631-12, 3631-24, 3631-36		46	3209,3210,3211,3212,3213,3214		28
Butyrometer stand	3330,3331	33	Cream butyrometer according to Roeder's method		
Butyrometer tubes for water baths WB 436	3717	54	3200,3201,3202,3203		28
Butyrometer tubes, closed	3766-G	54	Cream butyrometer according to Schulz-Kley's method	3208	28

Bezeichnung	Art.Nr.	Seite	Bezeichnung	Art.Nr.	Seite
CryoStar I freezing point determination device	7150	98	Glass plate for discharge viscometer	6521	81
CryoStar ^{automatic} freezing point determination device	7160	98	Glass stirrer	5430	72
Crystalline quartz sand	5460,5461,5462,5463,5464	73			
Culture cultivating appliance			Hanging cylinder	6820	85
	8610,8611,8612,8613,8614,8615,8616,8617	107	Hardware standardisation	3516	44
Culture tubes	9050,9054,9056	125	Head A for 36 buytrometer buckets	3685	52
Curd butyrometer 0-20%	3240	29	Head B for 8 Mojonniere tubes	3686	52
Cuvette chamber	8702	109	Head C for 6 solubility index tubes	3687	52
Cuvettes	8705	109	Heating cabinets	6000-6002,6008,6009	80
			Homogenisation pipette	3639	47
Dairy thermometer	7001,7031,7041	86	Hose pump for Lactoflash	3530-023	45
Delvotest	6570,6571	81	Hose pump for LactoStar	3510-023	44
Delvo-Test insert for water baths WB 436	3754	54	Humidity measuring device	5670	76
Desiccator	9201	126	Humidity/temperature measuring device	7115	88
Desiccator plate	9211	126	Hydrometer for amyl alcohol	6720	84
Digital second thermometer	7120	88	Hydrometer for boiler water	6690	84
Digital thermometer for contact-free measurements	7119	88	Hydrometer for buttermilk serum	6640,6641	83
Dilution flask	8290,8291	103	Hydrometer for condensed milk	6660-6661	83
Dilution pipettes	8300,8301,8302,8303	103	Hydrometer for milk	6620-6622	82
Dipper	3030,3031	9	Hydrometer for sulphuric acidter	6730,6731	84
Discharge viscometer	6520	81	Hydrometer for various liquids	6740-6743	84
Disposable plastic bag for BagMixer	3141	11	Hydrometer for yogurt and chocolate milk	6670	83
Double socke	9405,9406	129	Hydrometer for brine/Beaumé	6680,6681	84
Double-ended spatula	5440	72			
Drigalski spatula	8350	105	Ice cream and condensed milk butyrometer	3180,3181	27
Dry matter calculator	5360	72	Incubators	6035-6038	80
Durham tube	8130	102	Indicator papers	9365	128
			Infrared burner	5571	73
Erlenmeyer flasks	8817-8833	121	Inoculation wire	8370	73
Extraction tube according to Mojonniere's method	3870,3871	55	Insertion/immersion sensors	4350,4360,4361,4370	66
			Insertion/immersion sensors	7122,7213,7124,7125	89
Felt pen refill for contact pin	8504	106			
Filter bag for BagMixer	3142	11	Jolting volumeter	7660	100
Filter for reference drier	5704	77			
Filter papers	4910	71	Kapsenberg cap	8201	103
Filters, round	4911	71	KCl solution	4400	66
Flowmeter for reference drier	5708	78	Kjeldahl decomposition apparatus K8	4200	61
Foodstuff butyrometer according to Roeder's method	3250	29	Kjeldahl decomposition flask	4201	61
Free fat butyrometer	3252	29	Kjeldahl distillation apparatus S-3	4210	62
Freezing point determination	7150,716	98	Kjeldahl extraction station Behrosog 3	4203	61
Freezing point determination device CryoStar I	7150	98	Kjeldahl tabs	4230,4231	62
Freezing point determination device CryoStar ^{Automatic}	7160	98			
Frosted glass plate for bacteria counter ColonyStar	8502-001	106	Laboratory bottles	8970-8974	124
Funnels	8876-8879, 9235-9239	127	Laboratory clock	9440	129
			Laboratory furnaces	6220	81
Glass funnel	8876-8879	122	Laboratory lift	9300	127
Glass nail for powdered milk butyrometer	3315	31	Laboratory pH meter	4310,4311	65

Description	Art. no.	Page	Description	Art. no.	Page
Laboratory scales	5810,5811	79	Photometer	8700,8701	109
Laboratory vacuum pump	9470	129	Pipette helper	8920	123
Lactodensimeter according to Quevenne's method	6630, 6631	83	Pipette stand	3460	35
Lactodensimeter for milk	6600-6613	82	Pipette syringes	5110,5111,5112	71
LactoFlash	3530	45	Pipette tips for microlitre pipette	9510,9511	130
Lactometer	7500	100	Pipette tips for sampling pipette	7175	99
LactoStar	3510	36-38,44	Polyethylene funnels	9235-9239	127
Level gauge for lactodensimeters	6800	85	Polymeter	7110	88
Lid for Reductase test insert	3747	54	Powdered milk butyrometer acc. to Teichert's meth.	3170,3171	27
Lid for specimen receptacle	5702	77	Precision butyrometer	3150	26
Loading arm for reference drier filter	5705	78	Precision butyrometer for drinking milk and vat milk	3150	26
LOVIBOND comparator for resazurine tests	5160	72	Precision scales	5820	79
Low temperature laboratory thermometer	7081	87	Proportioning devices	9484,9485,9487,9488	130
Lyphan strips	9360-9364	128	Protein titration apparatus	4660	69
			Psychrometer	7100	88
M agnetic stirrer	8690,8691	108	Pump for LactoFlash	3530-023	45
Magnifying lens for bacteria counter ColonyStar	8501	106	Pump for LactoStar	3510-023	44
Mastitis detection	7920,7930,7931	101			
Matrix needle printer for humidity measuring device	5674	76	R eactivation solution	4422	67
Measuring cylinder	8850-8860	121,122	Reductase insert for water baths WG 436	3737	54
Measuring flask	8870-8875	122	Reference drier RD-8	5700	77
Measuring pipettes	8882-8887	12	Reference milk 1.5%, 3.5%, 30%, 0.10% fat		
Methylene blue tablets	5140	71		3517,3518,3519,3521	45
Microlitre pipettes	9495,9498	130	Reference table ADPI	7650	100
Microscope	8761,8762	110	Reference table with 3 purity grades	4920	71
Milk butyrometer	3151,3152,3153,3154,3155,3156,3157,3158	26	Refractometer (digital Abbe refractometer)	5620	75
Milk powder collector	3125	11	Refractometer (digital hand refractometer)	5614	75
Milk sample bottle 80 ml, 50 ml	3040, 3041	9,1	Refractometer (pocket refractometer)	5610,5612,5613	75
Milk sampler	3000,3001,3003,3004,3007,3008,3010,3011	9	Regfridgerated incubators	6070-6072	80
Milk stirrer	3021	9	Replacement butyrometer tube for NovaSafety	3641	47
Mixing cylinder	8862-8863	122	Replacement drive belt for solubility index mixer	7622	100
Mojonnier stand for water baths WB 436	3718	54	Replacement glass mixing bowl for solubility index mixer	7620	100
			Replacement measuring cylinder for jolting volumeter	7661	100
N arrow necked reagent bottles	8990-8995	124	Replacement stirrer for solubility index mixer	7621	100
Needle holder	8382	105	Replacement thermistor for CryoStar	7152	98
NovaSafety bench centrifuge	3670	47	Resazurine tablets	5150	71
			Rod thermometer, maximum-minimum	7095,7096	87
O utput regulator	5572	73	Rotation viscometer	6530	81
			Round aluminium foil	5712	78
P atent closure FIBU	3260	30	Rubber stopper for butyrometers	3280,3290,3300,3310	31
Patent closure GERBAL	3261	30	Rubber stopper for special solubility index tubes	3050	10
Patent closure NOVO	3262	30			
Pepsin hydrochloric acid solution	4421	67	S afety centrifuge	3680-L	52
Permanent automatic dispenser 3	390,3391	33	Safety goggles	3480	35
Peroxidase	7825	101	Safety reading light	3800	55
Petri dishes	8310,8312,8313,8314	103,104	Salt test	4760, 4770	69
pH meter	4310,4311,4315,4317,4450	65,66,67	Sample tube with marking at 2.0 ml	7167	99

Bezeichnung	Art.Nr.	Seite	Bezeichnung	Art.Nr.	Seite
Sampling pipette, adjustable	7174	99	Temperature sensor for Knick 911, 913	4319	66
Scales	5810,5811,5820	79	Temperature sensor for pH meter „pH 49“	4451	67
Scoop	3033,3034,3035	9	Test tube	8100	102
Sediment tester ASPILAC	4905	70	Test tube according to Beckel's method	5601	74
Sediment tester SEDILAB	4800,481	70	Test tube according to Beckel's method	5601	74
Separating funnel	5606	74	Test tube brush	9090	125
Separating funnel for extraction	5606	74	Test tube for LOVIBOND comparator	5162	72
Shaking hood	3340,3341	33	Test tube rack	9255-9256	127
Shaking machine for butyrometers	3852	55	Test tube shaking device	8650	108
Shaking machine for Mojonnier tubes	3850,3851	55	Test tubes	9080,9081	125
Shaking stand	3332	33	Thermal printer	7151	98
Shaking water bath	3550	45	Thermal printer for LactoStar, LactoFlash	7151	98
Single rod electrode	4452,4455	67	Thermistor for CryoStar	7152	98
Single rod electrode with temperature sensor	4380,4336	66	Thermometer	7001,7031,7041,7046,7060,7070, 7071,7081,7095,7096,7119,7120	86-88
Single rod electrode with temp. sensor for „pH49“	4453	67	Thin layer chromatography chamber	5607	74
Skim milk butyrometer according to Kehe's method	3161,3162	26	Thin layer chromatography plates	5608	74
Skim milk butyrometer acc. to Sichler's method	3160,3160-G	26	Titators	4654,4655	69
Skim milk butyrometer according to Siegfeld's method	3164	27	Titration apparatus SIMPLEX	4520,4521	68
Slide for microscope	8400	105	Titration apparatus STANDARD	4500,4501,4510	68
Smear needle	8340	105	Titration apparatus STANDARD	4500,4501,4510	68
Software for CryoStar	7156	98	Titration apparatus STANDARD	4500,4501,4510	68
Solubility index mixer	7610	100	Titration apparatus STANDARD	4500,4501,4510	68
Solubility index tubes	3634	46	Titration apparatus STANDARD	4500,4501,4510	68
Special solubility index tubes	36,37	46	Titration apparatus STANDARD	4500,4501,4510	68
Specimen receptacle for reference drier	5701	77	Titration apparatus STANDARD	4500,4501,4510	68
Staining cuvette	8430	106	Titration apparatus STANDARD	4500,4501,4510	68
Staining stand according to Bongert's method	8420	106	Titration apparatus STANDARD	4500,4501,4510	68
Stand clamp	9407,9408	129	Titration apparatus STANDARD	4500,4501,4510	68
Stand for 27 sample tubes	7168	99	Titration apparatus STANDARD	4500,4501,4510	68
Stand for bags for BagMixer	3144	11	Titration apparatus STANDARD	4500,4501,4510	68
Stand for extraction tubes	3875	55	Titration apparatus STANDARD	4500,4501,4510	68
Stand for lids and caps			Titration apparatus STANDARD	4500,4501,4510	68
for reference drier specimen receptacle	5707	78	Titration apparatus STANDARD	4500,4501,4510	68
Stand for permanent automatic dispenser	3400,3401,3402	33	Titration apparatus STANDARD	4500,4501,4510	68
Stand for special solubility index tubes	3330,3331	33	Titration apparatus STANDARD	4500,4501,4510	68
Stand ring with socketl	9409	129	Titration apparatus STANDARD	4500,4501,4510	68
Stand, tripod with cardanic suspension	6810,683	85	Titration apparatus STANDARD	4500,4501,4510	68
Stand for solubility index tubes	3636	46	Titration apparatus STANDARD	4500,4501,4510	68
Sterilisers	6047-6049	80	Titration apparatus STANDARD	4500,4501,4510	68
Sterilising box for petri dishes	8320	104	Titration apparatus STANDARD	4500,4501,4510	68
Sterilising box for pipettes	8190	102	Titration apparatus STANDARD	4500,4501,4510	68
Stirring rod	8696,8697,8698	109	Titration apparatus STANDARD	4500,4501,4510	68
Stop watch	6522	81	Titration apparatus STANDARD	4500,4501,4510	68
Stopper for extraction tubes acc. to Mojonnier's meth.	3872	55	Titration apparatus STANDARD	4500,4501,4510	68
Stopper with slit for milk sample bottles 50ml	3042	10	Titration apparatus STANDARD	4500,4501,4510	68
SuperVarioN, multi-purpose centrifuge	3680	48-52	Titration apparatus STANDARD	4500,4501,4510	68
Syringes for milk, cream	3440,3441,3442,3443,3450,3452	35	Titration apparatus STANDARD	4500,4501,4510	68
			Universal shelf for water baths WG 436	3727	54
			Universal thermometer	7046	87
			Viscosimeter	6520, 6530	81
			Volumetric pipettes for determination of fat content	3430-3438	34
			Volumetric pipettes	8888-8895	123
			Washbottles	9230-9233	126
			Water bath	8786,8788	111
			Water baths WB 436	3707,3708	53
			Water distillation apparatus	8771,8772	110
			Water paper	5600	74
			Weighing boat for butter	3322	32
			Weighing dish	5490	73
			Weighing dishes	9120,9121	125
			Weighing pipettes	3425,3426,3427,3428,3429	34
			Weighing stand for reference drier	5706	78
			Wide necked reagent bottles	8980-8985	124
			Wire basket	8543	107
			Wire cages for sterilisation	8330,8331,8332	104
			Wire cradle for milk samples	3091	10
			Wire mesh	8440,8441	106

